

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080624\
 Data File : VD079629.D
 Acq On : 06 Aug 2024 18:06
 Operator : RP/MD
 Sample : P3448-13RE
 Misc : 5.24G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP7RE

Quant Time: Aug 06 18:06:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	191275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	356802	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	272133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	81889	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	112621	57.957	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.920%
35) Dibromofluoromethane	7.804	113	120277	49.804	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.600%
50) Toluene-d8	10.269	98	377695	40.273	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	80.540%
62) 4-Bromofluorobenzene	12.575	95	74398	27.414	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	54.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080624\
Data File : VD079629.D
Acq On : 06 Aug 2024 18:06
Operator : RP/MD
Sample : P3448-13RE
Misc : 5.24G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP7RE

Quant Time: Aug 06 18:06:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

