

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080624\
 Data File : VD079631.D
 Acq On : 06 Aug 2024 19:01
 Operator : RP/MD
 Sample : P3374-23
 Misc : 7.63G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-862-K2-SO-13.5-14.0-072024

Quant Time: Aug 07 02:50:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 30 03:50:47 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	195568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	379469	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	329517	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	140913	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	101322	50.998	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.000%	
35) Dibromofluoromethane	7.805	113	114369	44.529	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 89.060%	
50) Toluene-d8	10.269	98	418806	41.989	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 83.980%	
62) 4-Bromofluorobenzene	12.569	95	105123	36.421	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 72.840%	
Target Compounds						
					Qvalue	
16) Acetone	4.034	43	1651	4.276	ug/l #	81
17) Carbon Disulfide	4.269	76	10061	1.442	ug/l	99
18) Methyl Acetate	4.563	43	7066	5.894	ug/l #	77
27) cis-1,2-Dichloroethene	7.075	96	13859	5.108	ug/l	99
44) Trichloroethene	9.028	130	55648	20.077	ug/l	91
58) 2-Chloroethyl Vinyl ether	9.881	63	2085	1.928	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080624\
Data File : VD079631.D
Acq On : 06 Aug 2024 19:01
Operator : RP/MD
Sample : P3374-23
Misc : 7.63G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-862-K2-SO-13.5-14.0-072024

Quant Time: Aug 07 02:50:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072924S.M
Quant Title : SW846 8260
QLast Update : Tue Jul 30 03:50:47 2024
Response via : Initial Calibration

