

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
 Data File : VD076940.D
 Acq On : 07 Aug 2023 19:27
 Operator : KP/SY
 Sample : 03903-03
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MW-03

Quant Time: Aug 08 04:41:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	219083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	435666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	427407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	192110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	111389	51.046	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.100%	
35) Dibromofluoromethane	7.804	113	145942	53.307	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 106.620%	
50) Toluene-d8	10.269	98	536512	54.780	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 109.560%	
62) 4-Bromofluorobenzene	12.575	95	170150	53.703	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 107.400%	
Target Compounds						
17) Carbon Disulfide	4.263	76	7888	1.560	ug/l	Qvalue # 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
Data File : VD076940.D
Acq On : 07 Aug 2023 19:27
Operator : KP/SY
Sample : 03903-03
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MW-03

Quant Time: Aug 08 04:41:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

