

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
 Data File : VD076942.D
 Acq On : 07 Aug 2023 20:24
 Operator : KP/SY
 Sample : 03903-05
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MW-05

Quant Time: Aug 08 04:42:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	158434	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	336060	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	338721	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	154187	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	100324	63.575	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	127.140%
35) Dibromofluoromethane	7.804	113	114959	54.436	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.880%
50) Toluene-d8	10.269	98	412111	54.550	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.100%
62) 4-Bromofluorobenzene	12.575	95	131769	53.916	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
Data File : VD076942.D
Acq On : 07 Aug 2023 20:24
Operator : KP/SY
Sample : 03903-05
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MW-05

Quant Time: Aug 08 04:42:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

