

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
 Data File : VD076929.D
 Acq On : 07 Aug 2023 14:19
 Operator : KP/SY
 Sample : VD0807SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBS01

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023

Quant Time: Aug 08 04:39:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	160653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	284877	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	264705	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	131427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	80650	50.401	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.800%			
35) Dibromofluoromethane	7.804	113	98751	55.162	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 110.320%			
50) Toluene-d8	10.269	98	325825	50.877	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 101.760%			
62) 4-Bromofluorobenzene	12.575	95	115677	55.835	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery = 111.680%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.934	85	22583	17.976	ug/l	99
3) Chloromethane	2.146	50	35370	16.321	ug/l	96
4) Vinyl Chloride	2.275	62	49460	17.121	ug/l	94
5) Bromomethane	2.669	94	43970	18.499	ug/l	100
6) Chloroethane	2.822	64	41301	18.936	ug/l	94
7) Trichlorofluoromethane	3.163	101	53137	20.672	ug/l	95
8) Diethyl Ether	3.593	74	15371	17.368	ug/l	85
9) 1,1,2-Trichlorotrifluo...	3.952	101	36460	21.578	ug/l	97
10) Methyl Iodide	4.157	142	31959	16.779	ug/l #	88
11) Tert butyl alcohol	5.057	59	12734	106.645	ug/l	97
12) 1,1-Dichloroethene	3.934	96	28656	18.051	ug/l	87
13) Acrolein	3.805	56	7584	93.203	ug/l #	62
14) Allyl chloride	4.552	41	36848	17.669	ug/l	94
15) Acrylonitrile	5.257	53	40745	93.736	ug/l	99
16) Acetone	4.022	43	36367	95.164	ug/l #	85
17) Carbon Disulfide	4.257	76	48644	13.119	ug/l	99
18) Methyl Acetate	4.569	43	28269	19.689	ug/l	92
19) Methyl tert-butyl Ether	5.310	73	78514	18.443	ug/l	96
20) Methylene Chloride	4.793	84	87603	35.514	ug/l	87
21) trans-1,2-Dichloroethene	5.310	96	32651	17.694	ug/l	91
22) Diisopropyl ether	6.216	45	96507	19.543	ug/l	93
23) Vinyl Acetate	6.157	43	214203m	87.667	ug/l	
24) 1,1-Dichloroethane	6.104	63	68499	20.573	ug/l	99
25) 2-Butanone	7.081	43	49862	97.491	ug/l	88
26) 2,2-Dichloropropane	7.075	77	59160	20.612	ug/l	98
27) cis-1,2-Dichloroethene	7.075	96	44582	19.598	ug/l	94
28) Bromochloromethane	7.434	49	17582	21.121	ug/l	91
29) Tetrahydrofuran	7.451	42	26914	89.065	ug/l	94
30) Chloroform	7.598	83	77645	21.291	ug/l	95
31) Cyclohexane	7.869	56	38841	16.493	ug/l	86
32) 1,1,1-Trichloroethane	7.793	97	63824	21.692	ug/l	99
36) 1,1-Dichloropropene	8.004	75	45832	19.592	ug/l	98
37) Ethyl Acetate	7.163	43	19238	18.389	ug/l	100
38) Carbon Tetrachloride	7.993	117	49221	21.975	ug/l	95
39) Methylcyclohexane	9.275	83	47375	17.539	ug/l #	89
40) Benzene	8.245	78	148207	20.079	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
 Data File : VD076929.D
 Acq On : 07 Aug 2023 14:19
 Operator : KP/SY
 Sample : VD0807SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBS01

Manual Integrations
 APPROVED

Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023

Quant Time: Aug 08 04:39:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.398	41	12668	19.830	ug/l	90
42) 1,2-Dichloroethane	8.322	62	42888	22.363	ug/l	97
43) Isopropyl Acetate	8.363	43	36899	18.073	ug/l #	82
44) Trichloroethene	9.028	130	40379	20.076	ug/l	88
45) 1,2-Dichloropropane	9.304	63	40255	21.128	ug/l	95
46) Dibromomethane	9.392	93	23170	20.776	ug/l	89
47) Bromodichloromethane	9.581	83	59503	21.261	ug/l	97
48) Methyl methacrylate	9.381	41	17325	18.644	ug/l	84
49) 1,4-Dioxane	9.392	88	5122	358.465	ug/l	95
51) 4-Methyl-2-Pentanone	10.163	43	105047	96.635	ug/l	93
52) Toluene	10.334	92	95228	19.644	ug/l	97
53) t-1,3-Dichloropropene	10.551	75	53452	20.577	ug/l	100
54) cis-1,3-Dichloropropene	10.022	75	62059	20.271	ug/l	94
55) 1,1,2-Trichloroethane	10.734	97	35623	21.874	ug/l	93
56) Ethyl methacrylate	10.598	69	35535	18.905	ug/l	92
57) 1,3-Dichloropropane	10.881	76	53595	20.208	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.869	63	72610	99.797	ug/l	99
59) 2-Hexanone	10.922	43	75861	97.802	ug/l	91
60) Dibromochloromethane	11.075	129	41351	21.857	ug/l	97
61) 1,2-Dibromoethane	11.181	107	31104	20.601	ug/l	97
64) Tetrachloroethene	10.810	164	30997	20.536	ug/l #	82
65) Chlorobenzene	11.604	112	112679	21.165	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.681	131	42475	21.430	ug/l	97
67) Ethyl Benzene	11.686	91	184891	19.775	ug/l	96
68) m/p-Xylenes	11.792	106	144298	40.186	ug/l	92
69) o-Xylene	12.122	106	69155	19.916	ug/l	91
70) Styrene	12.139	104	122520	20.156	ug/l	98
71) Bromoform	12.298	173	23937	22.051	ug/l #	94
73) Isopropylbenzene	12.422	105	184074	19.304	ug/l	95
74) N-amyl acetate	12.234	43	37993	18.913	ug/l	94
75) 1,1,2,2-Tetrachloroethane	12.675	83	42704	20.859	ug/l	97
76) 1,2,3-Trichloropropane	12.722	75	26938m	21.553	ug/l	
77) Bromobenzene	12.704	156	43460	19.746	ug/l	88
78) n-propylbenzene	12.763	91	230877	19.972	ug/l	96
79) 2-Chlorotoluene	12.851	91	123112	19.423	ug/l	98
80) 1,3,5-Trimethylbenzene	12.904	105	155076	20.113	ug/l	95
81) trans-1,4-Dichloro-2-b...	12.469	75	11757	20.020	ug/l	95
82) 4-Chlorotoluene	12.945	91	133553	20.169	ug/l	95
83) tert-Butylbenzene	13.169	119	134740	19.848	ug/l	98
84) 1,2,4-Trimethylbenzene	13.210	105	154415	19.598	ug/l	97
85) sec-Butylbenzene	13.345	105	210852	20.329	ug/l	96
86) p-Isopropyltoluene	13.463	119	170823	19.942	ug/l	96
87) 1,3-Dichlorobenzene	13.457	146	92661	20.468	ug/l	97
88) 1,4-Dichlorobenzene	13.539	146	94293	21.259	ug/l	98
89) n-Butylbenzene	13.786	91	165035	19.689	ug/l	97
90) Hexachloroethane	14.051	117	35238	22.251	ug/l	86
91) 1,2-Dichlorobenzene	13.833	146	82550	20.694	ug/l	97
92) 1,2-Dibromo-3-Chloropr...	14.451	75	4862	16.656	ug/l	83
93) 1,2,4-Trichlorobenzene	15.098	180	51000	20.942	ug/l	95
94) Hexachlorobutadiene	15.204	225	24755	21.135	ug/l	99
95) Naphthalene	15.333	128	98369	18.655	ug/l	100
96) 1,2,3-Trichlorobenzene	15.522	180	44899	20.545	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
Data File : VD076929.D
Acq On : 07 Aug 2023 14:19
Operator : KP/SY
Sample : VD0807SBSD01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023

Quant Time: Aug 08 04:39:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
Quant Title : SW846 8260
QLast Update : Mon Jul 17 08:07:52 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080723\
 Data File : VD076929.D
 Acq On : 07 Aug 2023 14:19
 Operator : KP/SY
 Sample : VD0807SBSD01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBSD01

Manual Integrations APPROVED

Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023


Quant Time: Aug 08 04:39:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D071423S.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 17 08:07:52 2023
 Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.875 min Scan# 1069
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion:168 Resp: 16065
Ion Ratio Lower Upper
168 100
99 51.5 36.7 55.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#2
Dichlorodifluoromethane
Concen: 17.976 ug/l
RT: 1.934 min Scan# 59
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion: 85 Resp: 22583
Ion Ratio Lower Upper
85 100
87 32.0 15.7 47.0





#3
Chloromethane
Concen: 16.321 ug/l
RT: 2.146 min Scan# 95
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 50 Resp: 35370
Ion Ratio Lower Upper
50 100
52 30.9 26.6 40.0




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#4
Vinyl Chloride
Concen: 17.121 ug/l
RT: 2.275 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion: 62 Resp: 49460
Ion Ratio Lower Upper
62 100
64 29.4 26.0 39.0





#5
Bromomethane
Concen: 18.499 ug/l
RT: 2.669 min Scan# 186
Delta R.T. -0.012 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion: 94 Resp: 43970
Ion Ratio Lower Upper
94 100
96 95.3 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#6
Chloroethane
Concen: 18.936 ug/l
RT: 2.822 min Scan# 210
Delta R.T. -0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Tgt Ion: 64 Resp: 41301
Ion Ratio Lower Upper
64 100
66 28.6 25.5 38.3





#7
 Trichlorofluoromethane
 Concen: 20.672 ug/l
 RT: 3.163 min Scan# 20
 Delta R.T. -0.001 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBSD01




Tgt Ion:101 Resp: 5313
 Ion Ratio Lower Upper
 101 100
 103 61.6 52.3 78.5

Manual Integrations
 APPROVED



Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023




#8
 Diethyl Ether
 Concen: 17.368 ug/l
 RT: 3.593 min Scan# 341
 Delta R.T. 0.006 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19



Tgt Ion: 74 Resp: 15371
 Ion Ratio Lower Upper
 74 100
 45 87.5 37.5 112.7









#9
1,1,2-Trichlorotrifluoroethane
Concen: 21.578 ug/l
RT: 3.952 min Scan# 403
Delta R.T. -0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBS01

Tgt Ion:101 Resp: 36460
Ion Ratio Lower Upper
101 100
85 42.2 33.8 50.6
151 70.9 59.8 89.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#10
Methyl Iodide
Concen: 16.779 ug/l
RT: 4.157 min Scan# 437
Delta R.T. -0.001 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion:142 Resp: 31959
Ion Ratio Lower Upper
142 100
127 45.1 29.4 44.0#
141 15.9 11.1 16.7









#11
Tert butyl alcohol
Concen: 106.645 ug/l
RT: 5.057 min Scan# 588
Delta R.T. 0.011 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBS01

Tgt Ion: 59 Resp: 12734
Ion Ratio Lower Upper
59 100
57 8.8 7.8 11.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#12
1,1-Dichloroethene
Concen: 18.051 ug/l
RT: 3.934 min Scan# 399
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 96 Resp: 28656
Ion Ratio Lower Upper
96 100
61 147.2 100.6 150.8
98 62.5 50.2 75.2





#13

Acrolein

Concen: 93.203 ug/l

RT: 3.805 min Scan# 31

Delta R.T. 0.012 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBSD01

Tgt Ion: 56 Resp: 7584

Ion Ratio Lower Upper

56 100





55 38.1 55.1 82.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#14

Allyl chloride

Concen: 17.669 ug/l

RT: 4.552 min Scan# 504

Delta R.T. -0.006 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19




Tgt Ion: 41 Resp: 36848


Ion Ratio Lower Upper

41 100

39 73.5 53.8 80.6


76 53.8 45.8 68.8





#15
Acrylonitrile
Concen: 93.736 ug/l
RT: 5.257 min Scan# 61
Delta R.T. 0.011 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 53 Resp: 40745
Ion Ratio Lower Upper
53 100
52 82.1 65.0 97.4
51 35.1 28.3 42.5




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#16
Acetone
Concen: 95.164 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 43 Resp: 36367
Ion Ratio Lower Upper
43 100
58 33.5 34.7 52.1#









#17
Carbon Disulfide
Concen: 13.119 ug/l
RT: 4.257 min Scan# 41
Delta R.T. -0.007 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 76 Resp: 4864
Ion Ratio Lower Upper
76 100
78 9.0 6.8 10.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#18
Methyl Acetate
Concen: 19.689 ug/l
RT: 4.569 min Scan# 507
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 43 Resp: 28269
Ion Ratio Lower Upper
43 100
74 28.7 26.4 39.6





#19

Methyl tert-butyl Ether

Concen: 18.443 ug/l

RT: 5.310 min Scan# 61

Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBS01

Tgt Ion: 73 Resp: 78514

Ion Ratio Lower Upper

73 100





57 22.7 16.6 25.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#20

Methylene Chloride

Concen: 35.514 ug/l

RT: 4.793 min Scan# 545

Delta R.T. -0.006 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion: 84 Resp: 87603




Ion Ratio Lower Upper


84 100

49 106.0 71.4 107.2

51 35.4 24.5 36.7


86 67.2 48.7 73.1





#21
trans-1,2-Dichloroethene
Concen: 17.694 ug/l
RT: 5.310 min Scan# 611
Delta R.T. 0.011 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBS01




Tgt Ion: 96 Resp: 3265
Ion Ratio Lower Upper
96 100
61 118.8 89.3 133.9
98 69.9 48.6 73.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#22
Diisopropyl ether
Concen: 19.543 ug/l
RT: 6.216 min Scan# 787
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Tgt Ion: 45 Resp: 96507
Ion Ratio Lower Upper
45 100
43 45.2 40.0 60.0
87 30.0 28.2 42.4
59 11.7 11.0 16.4





#23
 Vinyl Acetate
 Concen: 87.667 ug/l m
 RT: 6.157 min Scan# 777
 Delta R.T. 0.005 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19



Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBSD01



Tgt Ion: 43 Resp: 21420
 Ion Ratio Lower Upper
 43 100
 86 15.1 12.5 18.7




Manual Integrations
 APPROVED


Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023



#24
 1,1-Dichloroethane
 Concen: 20.573 ug/l
 RT: 6.104 min Scan# 768
 Delta R.T. -0.001 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19

Tgt Ion: 63 Resp: 68499
 Ion Ratio Lower Upper
 63 100
 98 9.4 4.5 13.7
 100 5.6 2.5 7.5









#25
2-Butanone
Concen: 97.491 ug/l
RT: 7.081 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 43 Resp: 4986
Ion Ratio Lower Upper
43 100
72 31.1 30.7 46.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#26
2,2-Dichloropropane
Concen: 20.612 ug/l
RT: 7.075 min Scan# 933
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 77 Resp: 59160
Ion Ratio Lower Upper
77 100
97 25.6 13.2 39.6









#27
cis-1,2-Dichloroethene
Concen: 19.598 ug/l
RT: 7.075 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 96 Resp: 4458
Ion Ratio Lower Upper
96 100
61 129.2 0.0 239.2
98 66.2 0.0 128.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#28
Bromochloromethane
Concen: 21.121 ug/l
RT: 7.434 min Scan# 994
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 49 Resp: 17582
Ion Ratio Lower Upper
49 100
129 1.8 0.0 6.0
130 94.4 83.2 124.8





#29

Tetrahydrofuran

Concen: 89.065 ug/l

RT: 7.451 min Scan# 995

Delta R.T. 0.011 min

Lab File: VD076929.D


Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBSD01






Tgt Ion:	42	Resp:	26914
Ion Ratio	100	Lower	Upper
42	100		
72	54.8	46.5	69.7
71	49.9	44.5	66.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#30

Chloroform

Concen: 21.291 ug/l




RT: 7.598 min Scan# 1022


Delta R.T. 0.005 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19


Tgt Ion:	83	Resp:	77645
Ion Ratio	100	Lower	Upper
83	100		
85	62.6	53.4	80.2





#31
Cyclohexane
Concen: 16.493 ug/l
RT: 7.869 min Scan# 1068
Delta R.T. -0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 56 Resp: 3884
Ion Ratio Lower Upper
56 100
69 40.1 28.6 43.0
84 122.8 84.8 127.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#32
1,1,1-Trichloroethane
Concen: 21.692 ug/l
RT: 7.793 min Scan# 1055
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 97 Resp: 63824
Ion Ratio Lower Upper
97 100
99 64.3 52.0 78.0
61 41.7 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 50.401 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBS01

Tgt Ion: 65 Resp: 80650

Ion Ratio Lower Upper

65 100





67 55.9 0.0 115.0

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.781 min Scan# 1223

Delta R.T. 0.006 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19




Tgt Ion:114 Resp: 284877


Ion Ratio Lower Upper

114 100

63 18.0 0.0 33.4

88 14.1 0.0 26.2





#35

Dibromofluoromethane

Concen: 55.162 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBS01





Tgt Ion: 113 Resp: 9875

Ion	Ratio	Lower	Upper
113	100		
111	105.6	84.7	127.1
192	18.0	16.8	25.2

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#36

1,1-Dichloropropene

Concen: 19.592 ug/l

RT: 8.004 min Scan# 1091




Delta R.T. -0.001 min


Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion: 75 Resp: 45832


Ion	Ratio	Lower	Upper
75	100		
110	38.0	19.9	59.6
77	29.7	24.7	37.1





#37
Ethyl Acetate
Concen: 18.389 ug/l
RT: 7.163 min Scan# 948
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion: 43 Resp: 19238
Ion Ratio Lower Upper
43 100
61 18.4 14.9 22.3
70 12.7 10.0 15.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#38
Carbon Tetrachloride
Concen: 21.975 ug/l
RT: 7.993 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19



Tgt Ion:117 Resp: 49221
Ion Ratio Lower Upper
117 100
119 102.5 78.0 117.0
121 31.7 24.2 36.2





#39

Methylcyclohexane

Concen: 17.539 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. -0.000 min

Lab File: VD076929.D


Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBS01






Tgt Ion: 83	Resp: 47375
Ion Ratio	Lower Upper
83	100
55	74.8 49.4 74.0
98	50.5 38.4 57.6

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#40

Benzene

Concen: 20.079 ug/l




RT: 8.245 min Scan# 1132


Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19


Tgt Ion: 78	Resp: 148207
Ion Ratio	Lower Upper
78	100
77	22.3 19.4 29.0





#41
Methacrylonitrile
Concen: 19.830 ug/l
RT: 7.398 min Scan# 91
Delta R.T. -0.001 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 41 Resp: 1266
Ion Ratio Lower Upper
41 100
39 50.4 34.0 51.0
67 86.7 77.0 115.4
52 29.4 27.9 41.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#42
1,2-Dichloroethane
Concen: 22.363 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 62 Resp: 42888
Ion Ratio Lower Upper
62 100
98 11.4 0.0 25.4





#43

Isopropyl Acetate

Concen: 18.073 ug/l

RT: 8.363 min Scan# 1151

Delta R.T. 0.006 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBS01

Tgt Ion: 43 Resp: 3689





Ion	Ratio	Lower	Upper
43	100		
61	23.2	31.4	47.0
87	18.3	14.9	22.3

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#44

Trichloroethene

Concen: 20.076 ug/l

RT: 9.028 min Scan# 1265




Delta R.T. -0.000 min


Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion:130 Resp: 40379

Ion	Ratio	Lower	Upper
130	100		
95	101.1	0.0	179.2





#45

1,2-Dichloropropane

Concen: 21.128 ug/l

RT: 9.304 min Scan# 1312

Delta R.T. 0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBSD01

Tgt Ion: 63 Resp: 4025

Ion Ratio Lower Upper

63 100





65 33.2 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023

Supervised By :Mahesh Dadoda 08/08/2023



#46

Dibromomethane

Concen: 20.776 ug/l

RT: 9.392 min Scan# 1327

Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19




Tgt Ion: 93 Resp: 23170


Ion Ratio Lower Upper

93 100

95 86.5 67.7 101.5


174 88.9 85.4 128.2





#47
Bromodichloromethane
Concen: 21.261 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion: 83 Resp: 5950
Ion Ratio Lower Upper
83 100
85 65.5 50.2 75.4
127 7.0 6.3 9.5

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#48
Methyl methacrylate
Concen: 18.644 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19



Tgt Ion: 41 Resp: 17325
Ion Ratio Lower Upper
41 100
69 106.3 102.2 153.2
39 59.0 41.4 62.2





#49

1,4-Dioxane

Concen: 358.465 ug/l

RT: 9.392 min Scan# 11

Delta R.T. 0.011 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :





VD0807SBSD01

Tgt Ion:	88	Resp:	512
Ion	Ratio	Lower	Upper
88	100		
43	26.4	19.0	28.6
58	57.8	43.5	65.3

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#50

Toluene-d8

Concen: 50.877 ug/l




RT: 10.269 min Scan# 1476


Delta R.T. -0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion:	98	Resp:	325825
Ion	Ratio	Lower	Upper
98	100		
100	65.0	50.3	75.5









#51
4-Methyl-2-Pentanone
Concen: 96.635 ug/l
RT: 10.163 min Scan# 1458
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 43 Resp: 10504
Ion Ratio Lower Upper
43 100
58 44.1 38.9 58.3




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#52
Toluene
Concen: 19.644 ug/l
RT: 10.334 min Scan# 1487
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion: 92 Resp: 95228
Ion Ratio Lower Upper
92 100
91 174.7 143.0 214.4





#53
t-1,3-Dichloropropene
Concen: 20.577 ug/l
RT: 10.551 min Scan# 11
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion: 75 Resp: 53452
Ion Ratio Lower Upper
75 100
77 31.5 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#54
cis-1,3-Dichloropropene
Concen: 20.271 ug/l
RT: 10.022 min Scan# 1434
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19



Tgt Ion: 75 Resp: 62059
Ion Ratio Lower Upper
75 100
77 27.9 25.5 38.3
39 30.3 22.1 33.1





#55

1,1,2-Trichloroethane

Concen: 21.874 ug/l

RT: 10.734 min Scan# 1555

Delta R.T. -0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :





VD0807SBSD01

Tgt Ion:	97	Resp:	3562
Ion Ratio	Lower	Upper	
97	100		
83	77.6	65.8	98.6
85	50.9	45.7	68.5
99	63.4	46.2	69.4

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#56

Ethyl methacrylate

Concen: 18.905 ug/l




RT: 10.598 min Scan# 1532


Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19


Tgt Ion:	69	Resp:	35535
Ion Ratio	Lower	Upper	
69	100		
41	54.8	39.1	58.7
39	27.9	20.0	30.0





#57
1,3-Dichloropropane
Concen: 20.208 ug/l
RT: 10.881 min Scan# 1580
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 76 Resp: 5359
Ion Ratio Lower Upper
76 100
78 31.5 26.4 39.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#58
2-Chloroethyl Vinyl ether
Concen: 99.797 ug/l
RT: 9.869 min Scan# 1408
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion: 63 Resp: 72610
Ion Ratio Lower Upper
63 100
106 31.9 25.3 37.9





#59
2-Hexanone
Concen: 97.802 ug/l
RT: 10.922 min Scan# 1587
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 43 Resp: 7586
Ion Ratio Lower Upper
43 100
58 60.8 34.1 102.2




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#60
Dibromochloromethane
Concen: 21.857 ug/l
RT: 11.075 min Scan# 1613
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion:129 Resp: 41351
Ion Ratio Lower Upper
129 100
127 76.8 37.1 111.5





#61
1,2-Dibromoethane
Concen: 20.601 ug/l
RT: 11.181 min Scan# 1631
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion:107 Resp: 31104
Ion Ratio Lower Upper
107 100
109 91.7 75.6 113.4

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#62
4-Bromofluorobenzene
Concen: 55.835 ug/l
RT: 12.575 min Scan# 1868
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19



Tgt Ion: 95 Resp: 115677
Ion Ratio Lower Upper
95 100
174 85.0 0.0 201.8
176 82.6 0.0 191.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.581 min Scan# 1

Delta R.T. -0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBSD01





Tgt Ion:117 Resp: 26470

Ion	Ratio	Lower	Upper
117	100		
82	48.0	38.3	57.5
119	31.8	25.0	37.4

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#64

Tetrachloroethene

Concen: 20.536 ug/l

RT: 10.810 min Scan# 1568




Delta R.T. 0.000 min


Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion:164 Resp: 30997


Ion	Ratio	Lower	Upper
164	100		
166	143.1	99.2	148.8
129	95.2	68.4	102.6
131	105.3	63.8	95.8





#65
Chlorobenzene
Concen: 21.165 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion:112 Resp: 112679
Ion Ratio Lower Upper
112 100
114 30.9 25.1 37.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 21.430 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion:131 Resp: 42475
Ion Ratio Lower Upper
131 100
133 94.0 48.5 145.5
119 69.0 35.6 106.9









#67
Ethyl Benzene
Concen: 19.775 ug/l
RT: 11.686 min Scan# 1716
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 91 Resp: 18489
Ion Ratio Lower Upper
91 100
106 31.0 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#68
m/p-Xylenes
Concen: 40.186 ug/l
RT: 11.792 min Scan# 1735
Delta R.T. -0.001 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion:106 Resp: 144298
Ion Ratio Lower Upper
106 100
91 194.9 147.0 220.6









#69
o-Xylene
Concen: 19.916 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBS01

Tgt Ion:106 Resp: 6915
Ion Ratio Lower Upper
106 100
91 209.2 97.9 293.6




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#70
Styrene
Concen: 20.156 ug/l
RT: 12.139 min Scan# 1794
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion:104 Resp: 122520
Ion Ratio Lower Upper
104 100
78 44.7 33.2 49.8
103 54.6 43.6 65.4





#71
Bromoform
Concen: 22.051 ug/l
RT: 12.298 min Scan# 1821
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion:173 Resp: 2393
Ion Ratio Lower Upper
173 100
175 45.4 24.6 73.8
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Tgt Ion:152 Resp: 131427
Ion Ratio Lower Upper
152 100
115 57.3 27.1 81.3
150 163.6 0.0 353.6





#73
Isopropylbenzene
Concen: 19.304 ug/l
RT: 12.422 min Scan# 184074
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion:105 Resp: 184074
Ion Ratio Lower Upper
105 100
120 26.0 14.3 42.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#74
N-ethyl acetate
Concen: 18.913 ug/l
RT: 12.234 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 43 Resp: 37993
Ion Ratio Lower Upper
43 100
70 48.7 45.0 67.4
55 28.1 23.8 35.6
61 27.4 22.1 33.1









#75
1,1,2,2-Tetrachloroethane
Concen: 20.859 ug/l
RT: 12.675 min Scan# 1885
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 83 Resp: 42704
Ion Ratio Lower Upper
83 100
131 8.7 5.1 15.3
85 63.0 32.6 97.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#76
1,2,3-Trichloropropane
Concen: 21.553 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion: 75 Resp: 26938
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 19.746 ug/l

RT: 12.704 min Scan# 1889

Delta R.T. 0.006 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :





VD0807SBS01

Tgt Ion:156 Resp: 43460
 Ion Ratio Lower Upper
 156 100
 77 177.6 76.8 230.4
 158 99.6 48.6 145.9

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023



#78

n-propylbenzene

Concen: 19.972 ug/l




RT: 12.763 min Scan# 1900


Delta R.T. -0.000 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion: 91 Resp: 230877
 Ion Ratio Lower Upper
 91 100
 120 22.6 12.4 37.2









#79
2-Chlorotoluene
Concen: 19.423 ug/l
RT: 12.851 min Scan# 1915
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion: 91 Resp: 123112
Ion Ratio Lower Upper
91 100
126 37.3 19.4 58.1




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#80
1,3,5-Trimethylbenzene
Concen: 20.113 ug/l
RT: 12.904 min Scan# 1924
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion:105 Resp: 155076
Ion Ratio Lower Upper
105 100
120 49.4 26.4 79.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.020 ug/l
RT: 12.469 min Scan# 1850
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01



Tgt Ion: 75 Resp: 1175
Ion Ratio Lower Upper
75 100
53 82.6 63.0 94.4
89 44.3 37.9 56.9




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#82
4-Chlorotoluene
Concen: 20.169 ug/l
RT: 12.945 min Scan# 1931
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19


Tgt Ion: 91 Resp: 133553
Ion Ratio Lower Upper
91 100
126 35.6 19.4 58.4





#83
tert-Butylbenzene
Concen: 19.848 ug/l
RT: 13.169 min Scan# 1969
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion:119 Resp: 134740
Ion Ratio Lower Upper
119 100
91 60.6 29.3 87.8
134 27.8 14.1 42.3

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#84
1,2,4-Trimethylbenzene
Concen: 19.598 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Tgt Ion:105 Resp: 154415
Ion Ratio Lower Upper
105 100
120 46.4 24.3 72.9





#85
 sec-Butylbenzene
 Concen: 20.329 ug/l
 RT: 13.345 min Scan# 1999
 Delta R.T. 0.000 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19




Instrument :
 MSVOA_D
 ClientSampleId :
 VD0807SBSD01




Tgt Ion:105 Resp: 210852
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.9 32.9

Manual Integrations
 APPROVED



Reviewed By :Romaben Patel 08/08/2023
 Supervised By :Mahesh Dadoda 08/08/2023




#86
 p-Isopropyltoluene
 Concen: 19.942 ug/l
 RT: 13.463 min Scan# 2019
 Delta R.T. 0.006 min
 Lab File: VD076929.D
 Acq: 07 Aug 2023 14:19



Tgt Ion:119 Resp: 170823
 Ion Ratio Lower Upper
 119 100
 134 26.1 14.2 42.6
 91 22.4 10.5 31.5









#87
1,3-Dichlorobenzene
Concen: 20.468 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion:146 Resp: 9266
Ion Ratio Lower Upper
146 100
111 39.4 17.9 53.8
148 63.9 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#88
1,4-Dichlorobenzene
Concen: 21.259 ug/l
RT: 13.539 min Scan# 2032
Delta R.T. -0.001 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion:146 Resp: 94293
Ion Ratio Lower Upper
146 100
111 37.2 17.4 52.3
148 62.8 31.7 95.1





#89

n-Butylbenzene

Concen: 19.689 ug/l

RT: 13.786 min Scan# 2074

Delta R.T. -0.001 min

Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Instrument :

MSVOA_D

ClientSampleId :

VD0807SBSD01





Tgt Ion: 91 Resp: 165038

Ion	Ratio	Lower	Upper
91	100		
92	51.6	26.5	79.5
134	24.7	13.8	41.3

Manual Integrations

APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#90

Hexachloroethane

Concen: 22.251 ug/l

RT: 14.051 min Scan# 2119




Delta R.T. 0.000 min


Lab File: VD076929.D

Acq: 07 Aug 2023 14:19

Tgt Ion:117 Resp: 35238


Ion	Ratio	Lower	Upper
117	100		
201	65.2	38.5	115.5





#91
1,2-Dichlorobenzene
Concen: 20.694 ug/l
RT: 13.833 min Scan# 2081
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion:146 Resp: 82550
Ion Ratio Lower Upper
146 100
111 39.6 18.3 54.8
148 63.9 31.6 94.7

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#92
1,2-Dibromo-3-Chloropropane
Concen: 16.656 ug/l
RT: 14.451 min Scan# 2187
Delta R.T. 0.006 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Tgt Ion: 75 Resp: 4862
Ion Ratio Lower Upper
75 100
155 110.9 47.9 143.8
157 140.7 59.9 179.7





#93
1,2,4-Trichlorobenzene
Concen: 20.942 ug/l
RT: 15.098 min Scan# 2297
Delta R.T. 0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19




Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01




Tgt Ion:180 Resp: 51000
Ion Ratio Lower Upper
180 100
182 91.8 48.5 145.6
145 29.6 14.2 42.6

Manual Integrations
APPROVED



Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023




#94
Hexachlorobutadiene
Concen: 21.135 ug/l
RT: 15.204 min Scan# 2315
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19



Tgt Ion:225 Resp: 24755
Ion Ratio Lower Upper
225 100
223 64.5 31.6 94.8
227 64.3 31.7 95.1









#95
Naphthalene
Concen: 18.655 ug/l
RT: 15.333 min Scan# 2337
Delta R.T. 0.005 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Instrument :
MSVOA_D
ClientSampleId :
VD0807SBSD01

Tgt Ion:128 Resp: 98369
Ion Ratio Lower Upper
128 100
127 13.0 10.5 15.7
129 10.8 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :Romaben Patel 08/08/2023
Supervised By :Mahesh Dadoda 08/08/2023



#96
1,2,3-Trichlorobenzene
Concen: 20.545 ug/l
RT: 15.522 min Scan# 2369
Delta R.T. -0.000 min
Lab File: VD076929.D
Acq: 07 Aug 2023 14:19

Tgt Ion:180 Resp: 44899
Ion Ratio Lower Upper
180 100
182 92.8 47.3 141.9
145 31.7 15.1 45.3

