

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080818\
 Data File : VD059712.D
 Acq On : 8 Aug 2018 14:46
 Operator : VA/AP
 Sample : J4297-04RE
 Misc : 5.52g/5ml/MSVOA D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-4RE

Manual Integrations
 APPROVED

apatel
 8/8/2018 4:51:37 PM

Quant Time: Aug 08 15:21:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 08 02:03:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	196876	50.00	ug/l	0.02
34) 1,4-Difluorobenzene	6.70	114	273707	50.00	ug/l	0.02
63) Chlorobenzene-d5	10.70	117	146623	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	12.92	152	82917	50.00	ug/l	0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	111205	43.91	ug/l	0.03
Spiked Amount 50.000			Recovery	=	87.82%	
35) Dibromofluoromethane	5.55	113	99907	40.64	ug/l	0.03
Spiked Amount 50.000			Recovery	=	81.28%	
50) Toluene-d8	8.71	98	194884	36.42	ug/l	0.02
Spiked Amount 50.000			Recovery	=	72.84%	
62) 4-Bromofluorobenzene	11.93	95	207823	84.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	168.30%	
Target Compounds						
16) Acetone	2.61	43	69860	146.85	ug/l	97
25) 2-Butanone	4.84	43	28648	45.85	ug/l	100
39) Methylcyclohexane	7.29	83	67270	26.16	ug/l	97
73) Isopropylbenzene	11.79	105	31857m	5.72	ug/l	
78) n-propylbenzene	12.17	91	26860m	3.89	ug/l	
85) sec-Butylbenzene	12.78	105	101498m	17.59	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080818\
Data File : VD059712.D
Acq On : 8 Aug 2018 14:46
Operator : VA/AP
Sample : J4297-04RE
Misc : 5.52g/5ml/MSVOA D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-4RE

Manual Integrations
APPROVED

apatel
8/8/2018 4:51:37 PM

Quant Time: Aug 08 15:21:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 08 02:03:19 2018
Response via : Initial Calibration

