

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD080818\
 Data File : VD059714.D
 Acq On : 8 Aug 2018 15:43
 Operator : VA/AP
 Sample : J4297-07RE
 Misc : 5.61g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-7RE

Quant Time: Aug 08 16:50:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 08 02:03:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	165032	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.69	114	209665	50.00	ug/l	0.01
63) Chlorobenzene-d5	10.70	117	95737	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	12.91	152	16542	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.01	65	94780	44.65	ug/l	0.01
Spiked Amount	50.000		Recovery	=	89.30%	
35) Dibromofluoromethane	5.55	113	79918	42.44	ug/l	0.02
Spiked Amount	50.000		Recovery	=	84.88%	
50) Toluene-d8	8.71	98	133871	32.66	ug/l	0.01
Spiked Amount	50.000		Recovery	=	65.32%	
62) 4-Bromofluorobenzene	11.91	95	57346	30.31	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	60.62%	
Target Compounds						
16) Acetone	2.60	43	12131	30.42	ug/l	97
31) Cyclohexane	5.59	56	20394	8.14	ug/l	92
39) Methylcyclohexane	7.28	83	88866	45.11	ug/l	97
52) Toluene	8.80	92	7325	2.54	ug/l	97
69) o-Xylene	11.40	106	3957	3.50	ug/l	66
83) tert-Butylbenzene	12.58	119	13294	12.83	ug/l	65
85) sec-Butylbenzene	12.76	105	24929	21.66	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD080818\
Data File : VD059714.D
Acq On : 8 Aug 2018 15:43
Operator : VA/AP
Sample : J4297-07RE
Misc : 5.61g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-7RE

Quant Time: Aug 08 16:50:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 08 02:03:19 2018
Response via : Initial Calibration

