


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD080818\
Data File : VD059706.D
Acq On : 8 Aug 2018 11:14
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
Title : SW846 8260
Last Update : Wed Aug 08 02:03:19 2018



AutoFind: Scans 1189, 1190, 1191; Background Corrected with Scan 1185

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.2	18271	PASS
75	95	30	60	54.9	39770	PASS
95	95	100	100	100.0	72413	PASS
96	95	5	9	6.7	4885	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	90.9	65850	PASS
175	174	5	9	8.2	5397	PASS
176	174	95	101	100.0	65834	PASS
177	176	5	9	6.3	4159	PASS