


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD080819\
Data File : VD063451.D
Acq On : 8 Aug 2019 21:29
Operator : JC/SY
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D072919S.M
Title : SW846 8260
Last Update : Tue Jul 30 03:04:50 2019



AutoFind: Scans 1232, 1233, 1234; Background Corrected with Scan 1227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	47465	PASS
75	95	30	60	43.4	97648	PASS
95	95	100	100	100.0	224834	PASS
96	95	5	9	5.7	12766	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	92.7	208485	PASS
175	174	5	9	7.2	15056	PASS
176	174	95	101	97.3	202818	PASS
177	176	5	9	5.9	11886	PASS