

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080823\
 Data File : VD076955.D
 Acq On : 08 Aug 2023 15:26
 Operator : KP/SY
 Sample : VIBLK
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VIBLK

Quant Time: Aug 09 05:44:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 09 05:34:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	179493	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	350559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	334713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	152447	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	88499	62.502	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.000%
35) Dibromofluoromethane	7.805	113	111233	59.020	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.040%
50) Toluene-d8	10.269	98	433905	71.681	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	143.360%#
62) 4-Bromofluorobenzene	12.575	95	126227	59.649	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	119.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080823\
Data File : VD076955.D
Acq On : 08 Aug 2023 15:26
Operator : KP/SY
Sample : VIBLK
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VIBLK

Quant Time: Aug 09 05:44:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 09 05:34:08 2023
Response via : Initial Calibration

