


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD080918\
Data File : VD059716.D
Acq On : 9 Aug 2018 13:33
Operator : VA/AP
Sample : BFB
Misc : 5.00g/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080618S.M
Title : SW846 8260
Last Update : Wed Aug 08 02:03:19 2018



AutoFind: Scans 1189, 1190, 1191; Background Corrected with Scan 1184

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.2	16954	PASS
75	95	30	60	55.4	38812	PASS
95	95	100	100	100.0	70026	PASS
96	95	5	9	6.7	4671	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	93.9	65720	PASS
175	174	5	9	8.1	5329	PASS
176	174	95	101	97.3	63957	PASS
177	176	5	9	6.1	3916	PASS