

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
 Data File : VD070007.D
 Acq On : 09 Aug 2021 20:18
 Operator : VA/SY
 Sample : M3299-03
 Misc : 6.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210439-25

Quant Time: Aug 10 07:06:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	504970	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1140528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1071036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	407440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	383159	65.975	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.960%
35) Dibromofluoromethane	7.914	113	379241	52.746	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.500%
50) Toluene-d8	10.338	98	1480249	53.393	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.626	95	446229	48.161	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	96.320%
Target Compounds						
17) Carbon Disulfide	4.409	76	78580	3.804	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
Data File : VD070007.D
Acq On : 09 Aug 2021 20:18
Operator : VA/SY
Sample : M3299-03
Misc : 6.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210439-25

Quant Time: Aug 10 07:06:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

