

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
 Data File : VD070009.D
 Acq On : 09 Aug 2021 21:16
 Operator : VA/SY
 Sample : M3318-06
 Misc : 5.69G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210433-19

Quant Time: Aug 10 07:07:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	565169	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1211415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1156964	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	489506	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	416393	64.061	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 128.120%			
35) Dibromofluoromethane	7.908	113	398751	52.215	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.420%			
50) Toluene-d8	10.338	98	1580869	53.685	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 107.380%			
62) 4-Bromofluorobenzene	12.626	95	510834	51.907	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 103.820%			
Target Compounds					Qvalue	
16) Acetone	4.156	43	144991	123.482	ug/l	98
17) Carbon Disulfide	4.409	76	98197	4.247	ug/l	97
25) 2-Butanone	7.220	43	14454	9.077	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
Data File : VD070009.D
Acq On : 09 Aug 2021 21:16
Operator : VA/SY
Sample : M3318-06
Misc : 5.69G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210433-19

Quant Time: Aug 10 07:07:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

