

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
 Data File : VD070014.D
 Acq On : 09 Aug 2021 23:40
 Operator : VA/SY
 Sample : M3328-01
 Misc : 9.11G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 20210457-43

Quant Time: Aug 10 07:09:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	729881	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1613506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1516133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	618130	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	505160	60.179	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.360%
35) Dibromofluoromethane	7.914	113	516204	50.750	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.500%
50) Toluene-d8	10.338	98	2095891	53.438	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.880%
62) 4-Bromofluorobenzene	12.626	95	685550	52.301	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.600%
Target Compounds						
16) Acetone	4.156	43	190929	125.910	ug/l	93
17) Carbon Disulfide	4.409	76	241175	8.077	ug/l	98
25) 2-Butanone	7.208	43	63129	30.699	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080921\
Data File : VD070014.D
Acq On : 09 Aug 2021 23:40
Operator : VA/SY
Sample : M3328-01
Misc : 9.11G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
20210457-43

Quant Time: Aug 10 07:09:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

