

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
 Data File : VD074001.D
 Acq On : 09 Aug 2022 21:27
 Operator : VA/SY
 Sample : VD0809SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0809SBL01

Quant Time: Aug 10 01:09:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 01:05:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	95275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	206287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	209077	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	94408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	68100	80.621	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 161.240%	
35) Dibromofluoromethane	7.908	113	78334	62.162	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 124.320%	
50) Toluene-d8	10.332	98	258713	69.500	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 139.000%	
62) 4-Bromofluorobenzene	12.620	95	90165	56.359	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 112.720%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.950	84	23565	15.941	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
Data File : VD074001.D
Acq On : 09 Aug 2022 21:27
Operator : VA/SY
Sample : VD0809SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0809SBL01

Quant Time: Aug 10 01:09:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 01:05:43 2022
Response via : Initial Calibration

