

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
 Data File : VD074007.D
 Acq On : 10 Aug 2022 00:56
 Operator : VA/SY
 Sample : N4046-07
 Misc : 6.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BEACONPOINT-BACKFILL-04

Quant Time: Aug 10 01:25:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 01:05:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	90803	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	204248	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	214352	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	98653	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	74296	92.288	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	184.580%#
35) Dibromofluoromethane	7.909	113	81009	64.927	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	129.860%
50) Toluene-d8	10.332	98	260056	70.558	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	141.120%#
62) 4-Bromofluorobenzene	12.620	95	90996	57.446	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	114.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
Data File : VD074007.D
Acq On : 10 Aug 2022 00:56
Operator : VA/SY
Sample : N4046-07
Misc : 6.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BEACONPOINT-BACKFILL-04

Quant Time: Aug 10 01:25:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 01:05:43 2022
Response via : Initial Calibration

