

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
 Data File : VD074009.D
 Acq On : 10 Aug 2022 01:55
 Operator : VA/SY
 Sample : N4046-11
 Misc : 7.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BEACONPOINT-BACKFILL-06

Quant Time: Aug 10 04:48:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 01:05:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	90624	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	200490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	214137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	100381	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	72530	90.273	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	180.540%#
35) Dibromofluoromethane	7.909	113	80155	65.447	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	130.900%
50) Toluene-d8	10.332	98	255445	70.606	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	141.220%#
62) 4-Bromofluorobenzene	12.620	95	89757	57.726	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	115.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
Data File : VD074009.D
Acq On : 10 Aug 2022 01:55
Operator : VA/SY
Sample : N4046-11
Misc : 7.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BEACONPOINT-BACKFILL-06

Quant Time: Aug 10 04:48:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 01:05:43 2022
Response via : Initial Calibration

