

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
 Data File : VD074014.D
 Acq On : 10 Aug 2022 04:22
 Operator : VA/SY
 Sample : N4046-21
 Misc : 6.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BEACONPOINT-BACKFILL-11

Quant Time: Aug 10 04:49:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 01:05:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	44037	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	96937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	108226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	52683	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	45848	117.431	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 234.860%#	
35) Dibromofluoromethane	7.903	113	43344	73.196	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 146.400%	
50) Toluene-d8	10.332	98	127619	72.956	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 145.920%#	
62) 4-Bromofluorobenzene	12.620	95	47056	62.592	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 125.180%	
Target Compounds						
20) Methylene Chloride	4.956	84	9068	12.223	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD080922\
Data File : VD074014.D
Acq On : 10 Aug 2022 04:22
Operator : VA/SY
Sample : N4046-21
Misc : 6.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BEACONPOINT-BACKFILL-11

Quant Time: Aug 10 04:49:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080922S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 01:05:43 2022
Response via : Initial Calibration

