

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081022\
 Data File : VD074028.D
 Acq On : 10 Aug 2022 16:35
 Operator : VA/SY
 Sample : N4038-06
 Misc : 4.22G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 S-6

Quant Time: Aug 10 17:41:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	123364	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	212118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	194998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	86711	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	56637	59.643	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.280%
35) Dibromofluoromethane	7.903	113	65989	46.838	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.680%
50) Toluene-d8	10.332	98	164378	43.910	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	87.820%
62) 4-Bromofluorobenzene	12.620	95	73907	42.658	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081022\
Data File : VD074028.D
Acq On : 10 Aug 2022 16:35
Operator : VA/SY
Sample : N4038-06
Misc : 4.22G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
S-6

Quant Time: Aug 10 17:41:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

