



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081022\
Data File : VD074016.D
Acq On : 10 Aug 2022 09:10
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Title : SW846 8260
Last Update : Wed Aug 10 14:55:58 2022



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.7	8575	PASS
75	95	30	60	57.1	23707	PASS
95	95	100	100	100.0	41485	PASS
96	95	5	9	6.9	2880	PASS
173	174	0.00	2	1.6	553	PASS
174	95	50	100	82.7	34315	PASS
175	174	5	9	8.3	2832	PASS
176	174	95	101	95.3	32704	PASS
177	176	5	9	6.5	2115	PASS