

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081023\
 Data File : VD076970.D
 Acq On : 10 Aug 2023 18:49
 Operator : JC/SY
 Sample : 03965-14
 Misc : 5.40G/10.0ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A48K6

Quant Time: Aug 11 03:07:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM080323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Aug 11 02:51:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	264032	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	259820	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	119152	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	140202	19.309	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	77.240%
7) Chloroethane-d5	2.793	69	138662	20.657	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.640%
11) 1,1-Dichloroethene-d2	3.905	65	32278	18.761	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.040%
21) 2-Butanone-d5	6.981	46	43304	37.146	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	74.300%
24) Chloroform-d	7.563	84	174428	20.288	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.160%
26) 1,2-Dichloroethane-d4	8.228	65	88271	20.595	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	82.400%
32) Benzene-d6	8.199	84	351177	21.498	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	86.000%
36) 1,2-Dichloropropane-d6	9.210	67	107074	21.319	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	85.280%
41) Toluene-d8	10.269	98	325856	21.757	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.040%
43) trans-1,3-Dichloroprop...	10.522	79	45143	19.880	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.520%
47) 2-Hexanone-d5	10.881	63	37673	39.267	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.540%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	95476	19.673	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.680%
66) 1,2-Dichlorobenzene-d4	13.816	152	103872	23.845	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.400%
Target Compounds					Qvalue	
13) Acetone	4.017	43	10968	9.201	ug/L	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081023\
Data File : VD076970.D
Acq On : 10 Aug 2023 18:49
Operator : JC/SY
Sample : 03965-14
Misc : 5.40G/10.0ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A48K6

Quant Time: Aug 11 03:07:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM080323SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Aug 11 02:51:25 2023
Response via : Initial Calibration

