

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070017.D
 Acq On : 11 Aug 2021 12:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 12 05:27:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	557528	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	962805	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	893447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	421947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	295187	46.036	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.080%
35) Dibromofluoromethane	7.914	113	300460	49.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.000%
50) Toluene-d8	10.338	98	1164636	49.763	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.520%
62) 4-Bromofluorobenzene	12.626	95	405106	51.793	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.580%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.985	85	238024	40.561	ug/l	99
3) Chloromethane	2.209	50	344529	40.473	ug/l	97
4) Vinyl Chloride	2.344	62	397596	44.407	ug/l	98
5) Bromomethane	2.756	94	247486	46.729	ug/l	99
6) Chloroethane	2.915	64	264411	45.562	ug/l	99
7) Trichlorofluoromethane	3.268	101	518158	46.880	ug/l	98
8) Diethyl Ether	3.709	74	154900	49.823	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.091	101	341104	48.592	ug/l	98
10) Methyl Iodide	4.297	142	299744	43.716	ug/l	99
11) Tert butyl alcohol	5.220	59	83852	252.727	ug/l	99
12) 1,1-Dichloroethene	4.062	96	305892	47.954	ug/l	97
13) Acrolein	3.921	56	84398	181.715	ug/l	99
14) Allyl chloride	4.715	41	479151	51.233	ug/l	98
15) Acrylonitrile	5.415	53	375566	256.795	ug/l	100
16) Acetone	4.156	43	346615	299.240	ug/l	99
17) Carbon Disulfide	4.403	76	958073	42.006	ug/l	99
18) Methyl Acetate	4.720	43	164278	48.880	ug/l	96
19) Methyl tert-butyl Ether	5.473	73	706525	53.458	ug/l	100
20) Methylene Chloride	4.962	84	398023	47.073	ug/l	97
21) trans-1,2-Dichloroethene	5.467	96	359424	47.818	ug/l	95
22) Diisopropyl ether	6.362	45	1090330	53.651	ug/l	98
23) Vinyl Acetate	6.303	43	2745515	245.946	ug/l	99
24) 1,1-Dichloroethane	6.262	63	681110	48.523	ug/l	99
25) 2-Butanone	7.209	43	435379	277.167	ug/l	97
26) 2,2-Dichloropropane	7.209	77	565070	52.941	ug/l	99
27) cis-1,2-Dichloroethene	7.209	96	396431	49.427	ug/l	99
28) Bromochloromethane	7.544	49	280086	52.555	ug/l	98
29) Tetrahydrofuran	7.561	42	278064	273.612	ug/l	99
30) Chloroform	7.709	83	663539	47.017	ug/l	92
31) Cyclohexane	7.985	56	608963	48.606	ug/l	98
32) 1,1,1-Trichloroethane	7.903	97	575256	49.245	ug/l	100
36) 1,1-Dichloropropene	8.109	75	526907	50.712	ug/l	100
37) Ethyl Acetate	7.291	43	203568	53.818	ug/l	100
38) Carbon Tetrachloride	8.091	117	486760	51.287	ug/l	98
39) Methylcyclohexane	9.355	83	634963	53.359	ug/l	97
40) Benzene	8.350	78	1498925	49.796	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070017.D
 Acq On : 11 Aug 2021 12:15
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050

Quant Time: Aug 12 05:27:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.520	41	104625	52.707	ug/l	99
42) 1,2-Dichloroethane	8.420	62	379121	47.624	ug/l	98
43) Isopropyl Acetate	8.450	43	365830	53.926	ug/l	98
44) Trichloroethene	9.108	130	362913	48.957	ug/l	99
45) 1,2-Dichloropropane	9.385	63	390883	49.437	ug/l	100
46) Dibromomethane	9.467	93	192183	49.448	ug/l	99
47) Bromodichloromethane	9.661	83	499930	49.509	ug/l	95
48) Methyl methacrylate	9.450	41	182740	50.559	ug/l	98
49) 1,4-Dioxane	9.456	88	43627	1062.471	ug/l	98
51) 4-Methyl-2-Pentanone	10.226	43	1003278	284.199	ug/l	99
52) Toluene	10.403	92	947859	51.909	ug/l	99
53) t-1,3-Dichloropropene	10.614	75	446137	52.636	ug/l	99
54) cis-1,3-Dichloropropene	10.085	75	534838	52.121	ug/l	99
55) 1,1,2-Trichloroethane	10.797	97	272043	49.557	ug/l	97
56) Ethyl methacrylate	10.655	69	335407	49.463	ug/l	98
57) 1,3-Dichloropropane	10.938	76	473424	50.966	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.938	63	793545	264.501	ug/l	99
59) 2-Hexanone	10.979	43	702413	300.811	ug/l	99
60) Dibromochloromethane	11.132	129	316193	50.859	ug/l	99
61) 1,2-Dibromoethane	11.244	107	248977	50.375	ug/l	99
64) Tetrachloroethene	10.873	164	284402	49.937	ug/l	97
65) Chlorobenzene	11.667	112	955650	51.628	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.738	131	342301	51.405	ug/l	99
67) Ethyl Benzene	11.738	91	1781920	54.679	ug/l	99
68) m/p-Xylenes	11.844	106	1389518	109.989	ug/l	100
69) o-Xylene	12.173	106	626873	54.582	ug/l	99
70) Styrene	12.185	104	1106368	55.586	ug/l	100
71) Bromoform	12.349	173	168143	51.995	ug/l #	100
73) Isopropylbenzene	12.473	105	1700555	56.092	ug/l	100
74) N-amyl acetate	12.279	43	361812	58.020	ug/l	99
75) 1,1,2,2-Tetrachloroethane	12.720	83	312024	51.536	ug/l	99
76) 1,2,3-Trichloropropane	12.773	75	397069	88.533	ug/l #	1
77) Bromobenzene	12.749	156	348738	51.073	ug/l	98
78) n-propylbenzene	12.808	91	2190013	56.242	ug/l	100
79) 2-Chlorotoluene	12.897	91	1215473	53.739	ug/l	100
80) 1,3,5-Trimethylbenzene	12.949	105	1446444	55.773	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.520	75	86033	58.194	ug/l	97
82) 4-Chlorotoluene	12.996	91	1255772	53.107	ug/l	100
83) tert-Butylbenzene	13.214	119	1196061	56.724	ug/l	99
84) 1,2,4-Trimethylbenzene	13.261	105	1424464	55.675	ug/l	98
85) sec-Butylbenzene	13.391	105	1908960	57.268	ug/l	99
86) p-Isopropyltoluene	13.508	119	1540418	56.881	ug/l	99
87) 1,3-Dichlorobenzene	13.508	146	737537	51.447	ug/l	100
88) 1,4-Dichlorobenzene	13.585	146	715794	50.095	ug/l	99
89) n-Butylbenzene	13.832	91	1525488	57.078	ug/l	99
90) Hexachloroethane	14.102	117	304859	53.143	ug/l	99
91) 1,2-Dichlorobenzene	13.879	146	644093	52.055	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.496	75	44523	51.068	ug/l	97
93) 1,2,4-Trichlorobenzene	15.149	180	363198	54.405	ug/l	99
94) Hexachlorobutadiene	15.255	225	218482	53.501	ug/l	99
95) Naphthalene	15.385	128	688721	49.944	ug/l	100
96) 1,2,3-Trichlorobenzene	15.579	180	314711	54.512	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070017.D
Acq On : 11 Aug 2021 12:15
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1


Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Quant Time: Aug 12 05:27:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 12 05:27:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:168 Resp: 557528
Ion Ratio Lower Upper
168 100
99 62.1 51.1 76.7



#2
Dichlorodifluoromethane
Concen: 40.561 ug/l
RT: 1.985 min Scan# 11
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 85 Resp: 238024
Ion Ratio Lower Upper
85 100
87 32.8 16.1 48.3









#3
Chloromethane
Concen: 40.473 ug/l
RT: 2.209 min Scan# 49
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 50 Resp: 344529
Ion Ratio Lower Upper
50 100
52 33.6 25.7 38.5



#4
Vinyl Chloride
Concen: 44.407 ug/l
RT: 2.344 min Scan# 72
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 62 Resp: 397596
Ion Ratio Lower Upper
62 100
64 32.5 25.0 37.4









#5
Bromomethane
Concen: 46.729 ug/l
RT: 2.756 min Scan# 142
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 247486
Ion Ratio Lower Upper
94 100
96 92.1 74.8 112.2



#6
Chloroethane
Concen: 45.562 ug/l
RT: 2.915 min Scan# 169
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 64 Resp: 264411
Ion Ratio Lower Upper
64 100
66 34.2 27.8 41.6









#7
Trichlorofluoromethane
Concen: 46.880 ug/l
RT: 3.268 min Scan# 229
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 518158
Ion Ratio Lower Upper
101 100
103 65.7 51.3 76.9



#8
Diethyl Ether
Concen: 49.823 ug/l
RT: 3.709 min Scan# 304
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 74 Resp: 154900
Ion Ratio Lower Upper
74 100
45 95.9 46.8 140.4









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.592 ug/l
RT: 4.091 min Scan# 369
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 341104
Ion Ratio Lower Upper
101 100
85 42.7 34.7 52.1
151 67.6 55.4 83.0



#10
Methyl Iodide
Concen: 43.716 ug/l
RT: 4.297 min Scan# 404
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15


Tgt Ion:142 Resp: 299744
Ion Ratio Lower Upper
142 100
127 44.2 35.7 53.5
141 14.6 12.0 18.0








#11
Tert butyl alcohol
Concen: 252.727 ug/l
RT: 5.220 min Scan# 561
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050







Tgt Ion: 59 Resp: 83852
Ion Ratio Lower Upper
59 100
57 9.9 8.1 12.1



#12
1,1-Dichloroethene
Concen: 47.954 ug/l
RT: 4.062 min Scan# 364
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 96 Resp: 305892
Ion Ratio Lower Upper
96 100
61 161.2 126.9 190.3
98 62.1 53.2 79.8









#13
Acrolein
Concen: 181.715 ug/l
RT: 3.921 min Scan# 340
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 84398
Ion Ratio Lower Upper
56 100
55 70.4 57.2 85.8



#14
Allyl chloride
Concen: 51.233 ug/l
RT: 4.715 min Scan# 475
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 41 Resp: 479151
Ion Ratio Lower Upper
41 100
39 61.0 49.8 74.6
76 40.7 31.8 47.8









#15
Acrylonitrile
Concen: 256.795 ug/l
RT: 5.415 min Scan# 594
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 375566
Ion Ratio Lower Upper
53 100
52 81.5 65.3 97.9
51 35.5 28.3 42.5



#16
Acetone
Concen: 299.240 ug/l
RT: 4.156 min Scan# 380
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 43 Resp: 346615
Ion Ratio Lower Upper
43 100
58 36.3 29.4 44.0









#17
Carbon Disulfide
Concen: 42.006 ug/l
RT: 4.403 min Scan# 422
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 958073
Ion Ratio Lower Upper
76 100
78 9.4 7.3 10.9



#18
Methyl Acetate
Concen: 48.880 ug/l
RT: 4.720 min Scan# 476
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 43 Resp: 164278
Ion Ratio Lower Upper
43 100
74 24.7 21.4 32.2









#19
Methyl tert-butyl Ether
Concen: 53.458 ug/l
RT: 5.473 min Scan# 604
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 73 Resp: 706525
Ion Ratio Lower Upper
73 100
57 22.2 17.8 26.8



#20
Methylene Chloride
Concen: 47.073 ug/l
RT: 4.962 min Scan# 517
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15


Tgt Ion: 84 Resp: 398023
Ion Ratio Lower Upper
84 100
49 119.2 92.4 138.6
51 38.0 28.5 42.7
86 60.4 48.2 72.4








#21
trans-1,2-Dichloroethene
Concen: 47.818 ug/l
RT: 5.467 min Scan# 603
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15


Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050





Tgt Ion: 96 Resp: 359424
Ion Ratio Lower Upper
96 100
61 141.3 107.0 160.6
98 63.2 52.3 78.5




#22
Diisopropyl ether
Concen: 53.651 ug/l
RT: 6.362 min Scan# 755
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15



Tgt Ion: 45 Resp: 1090330
Ion Ratio Lower Upper
45 100
43 51.2 39.2 58.8
87 27.7 22.3 33.5
59 12.7 9.4 14.0









#23
 Vinyl Acetate
 Concen: 245.946 ug/l
 RT: 6.303 min Scan# 745
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 2745515
 Ion Ratio Lower Upper
 43 100
 86 11.1 8.6 13.0



#24
 1,1-Dichloroethane
 Concen: 48.523 ug/l
 RT: 6.262 min Scan# 738
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15

Tgt Ion: 63 Resp: 681110
 Ion Ratio Lower Upper
 63 100
 98 6.8 3.6 10.8
 100 4.2 2.2 6.6









#25
2-Butanone
Concen: 277.167 ug/l
RT: 7.209 min Scan# 899
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 435379
Ion Ratio Lower Upper
43 100
72 30.2 22.8 34.2



#26
2,2-Dichloropropane
Concen: 52.941 ug/l
RT: 7.209 min Scan# 899
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 77 Resp: 565070
Ion Ratio Lower Upper
77 100
97 23.4 11.9 35.7









#27
 cis-1,2-Dichloroethene
 Concen: 49.427 ug/l
 RT: 7.209 min Scan# 899
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050

Tgt Ion: 96 Resp: 396431




Ion	Ratio	Lower	Upper
96	100		
61	147.6	0.0	291.6
98	65.6	0.0	132.0




#28
 Bromochloromethane
 Concen: 52.555 ug/l
 RT: 7.544 min Scan# 956
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15

Tgt Ion: 49 Resp: 280086

Ion	Ratio	Lower	Upper
49	100		
129	1.8	0.0	4.2
130	72.8	59.6	89.4





#29

Tetrahydrofuran

Concen: 273.612 ug/l

RT: 7.561 min Scan# 959





Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 42 Resp: 278064

Ion	Ratio	Lower	Upper
42	100		
72	46.2	36.5	54.7
71	43.0	33.5	50.3



#30

Chloroform

Concen: 47.017 ug/l

RT: 7.709 min Scan# 984




Delta R.T. -0.000 min


Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 83 Resp: 663539

Ion	Ratio	Lower	Upper
83	100		
85	67.7	49.1	73.7





#31

Cyclohexane

Concen: 48.606 ug/l

RT: 7.985 min Scan# 1031

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050





Tgt Ion: 56 Resp: 608963

Ion Ratio Lower Upper

56 100

69 33.7 27.9 41.9

84 87.2 71.4 107.2



#32

1,1,1-Trichloroethane

Concen: 49.245 ug/l

RT: 7.903 min Scan# 1017

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15




Tgt Ion: 97 Resp: 575256


Ion Ratio Lower Upper

97 100

99 64.0 50.8 76.2

61 43.3 34.5 51.7









#33
1,2-Dichloroethane-d4
Concen: 46.036 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 65 Resp: 295187
Ion Ratio Lower Upper
65 100
67 53.1 0.0 111.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 114 Resp: 962805
Ion Ratio Lower Upper
114 100
63 19.8 0.0 41.6
88 16.0 0.0 31.4





#35

Dibromofluoromethane

Concen: 49.503 ug/l

RT: 7.914 min Scan# 1019

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15





Tgt Ion:113 Resp: 300460

Ion Ratio Lower Upper

113 100

111 102.3 80.5 120.7

192 16.4 12.6 19.0



#36

1,1-Dichloropropene

Concen: 50.712 ug/l

RT: 8.109 min Scan# 1052

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15




Tgt Ion: 75 Resp: 526907


Ion Ratio Lower Upper

75 100

110 33.4 16.9 50.7

77 31.2 24.9 37.3





#37

Ethyl Acetate

Concen: 53.818 ug/l

RT: 7.291 min Scan# 913

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Instrument :





MSVOA_D

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 203568

Ion	Ratio	Lower	Upper
43	100		
61	14.5	11.8	17.8
70	11.1	8.9	13.3



#38

Carbon Tetrachloride

Concen: 51.287 ug/l

RT: 8.091 min Scan# 1049




Delta R.T. -0.006 min


Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 117 Resp: 486760

Ion	Ratio	Lower	Upper
117	100		
119	97.7	76.2	114.2
121	30.4	24.5	36.7





#39

Methylcyclohexane

Concen: 53.359 ug/l

RT: 9.355 min Scan# 1264

Delta R.T. 0.006 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 83 Resp: 634963





Ion	Ratio	Lower	Upper
83	100		
55	75.0	62.2	93.2
98	48.4	37.8	56.8

Instrument :

MSVOA_D

ClientSampleId :

VSTDCCC050



#40

Benzene

Concen: 49.796 ug/l

RT: 8.350 min Scan# 1093




Delta R.T. -0.000 min


Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 78 Resp: 1498925


Ion	Ratio	Lower	Upper
78	100		
77	23.3	19.4	29.0








#41
Methacrylonitrile
Concen: 52.707 ug/l
RT: 7.520 min Scan# 952
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15


Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050





Tgt Ion:	41	Resp:	104625
Ion Ratio	100	Lower	Upper
39	57.8	46.7	70.1
67	86.1	69.3	103.9
52	33.8	28.2	42.4




#42
1,2-Dichloroethane
Concen: 47.624 ug/l
RT: 8.420 min Scan# 1105
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15



Tgt Ion:	62	Resp:	379121
Ion Ratio	100	Lower	Upper
98	10.9	0.0	20.4









#43
Isopropyl Acetate
Concen: 53.926 ug/l
RT: 8.450 min Scan# 1110
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 365830
Ion Ratio Lower Upper
43 100
61 33.6 27.8 41.6
87 14.6 11.3 16.9



#44
Trichloroethene
Concen: 48.957 ug/l
RT: 9.108 min Scan# 1222
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:130 Resp: 362913
Ion Ratio Lower Upper
130 100
95 104.0 0.0 210.8









#45

1,2-Dichloropropane
Concen: 49.437 ug/l
RT: 9.385 min Scan# 1269
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050




Tgt Ion: 63 Resp: 390883
Ion Ratio Lower Upper
63 100
65 31.3 24.9 37.3




#46

Dibromomethane
Concen: 49.448 ug/l
RT: 9.467 min Scan# 1283
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 93 Resp: 192183
Ion Ratio Lower Upper
93 100
95 82.9 67.5 101.3
174 84.1 66.3 99.5









#47

Bromodichloromethane
Concen: 49.509 ug/l
RT: 9.661 min Scan# 1316
Delta R.T. 0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 499930
Ion Ratio Lower Upper
83 100
85 61.0 52.0 78.0
127 8.9 6.6 10.0




#48

Methyl methacrylate
Concen: 50.559 ug/l
RT: 9.450 min Scan# 1280
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 41 Resp: 182740
Ion Ratio Lower Upper
41 100
69 89.6 73.4 110.2
39 45.1 36.6 55.0









#49
1,4-Dioxane
Concen: 1062.471 ug/l
RT: 9.456 min Scan# 1281
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 88 Resp: 43627
Ion Ratio Lower Upper
88 100
43 30.4 26.8 40.2
58 75.0 59.8 89.6



#50
Toluene-d8
Concen: 49.763 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 98 Resp: 1164636
Ion Ratio Lower Upper
98 100
100 65.5 51.9 77.9









#51
4-Methyl-2-Pentanone
Concen: 284.199 ug/l
RT: 10.226 min Scan# 1412
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 1003278
Ion Ratio Lower Upper
43 100
58 41.7 33.7 50.5



#52
Toluene
Concen: 51.909 ug/l
RT: 10.403 min Scan# 1442
Delta R.T. 0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 92 Resp: 947859
Ion Ratio Lower Upper
92 100
91 172.5 138.7 208.1









#53
t-1,3-Dichloropropene
Concen: 52.636 ug/l
RT: 10.614 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 446137
Ion Ratio Lower Upper
75 100
77 31.8 25.8 38.6



#54
cis-1,3-Dichloropropene
Concen: 52.121 ug/l
RT: 10.085 min Scan# 1388
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 75 Resp: 534838
Ion Ratio Lower Upper
75 100
77 31.3 26.3 39.5
39 38.7 31.0 46.6









#55
1,1,2-Trichloroethane
Concen: 49.557 ug/l
RT: 10.797 min Scan# 1509
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:	97	Resp:	272043
Ion	Ratio	Lower	Upper
97	100		
83	82.7	69.1	103.7
85	56.3	46.2	69.2
99	59.9	48.8	73.2



#56
Ethyl methacrylate
Concen: 49.463 ug/l
RT: 10.655 min Scan# 1485
Delta R.T. 0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:	69	Resp:	335407
Ion	Ratio	Lower	Upper
69	100		
41	60.3	50.2	75.2
39	27.1	22.0	33.0









#57
1,3-Dichloropropane
Concen: 50.966 ug/l
RT: 10.938 min Scan# 1533
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 473424
Ion Ratio Lower Upper
76 100
78 32.8 26.2 39.2



#58
2-Chloroethyl Vinyl ether
Concen: 264.501 ug/l
RT: 9.938 min Scan# 1363
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 63 Resp: 793545
Ion Ratio Lower Upper
63 100
106 26.6 21.8 32.6









#59
2-Hexanone
Concen: 300.811 ug/l
RT: 10.979 min Scan# 1540
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 43 Resp: 702413
Ion Ratio Lower Upper
43 100
58 59.8 29.4 88.3



#60
Dibromochloromethane
Concen: 50.859 ug/l
RT: 11.132 min Scan# 1566
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 129 Resp: 316193
Ion Ratio Lower Upper
129 100
127 76.7 38.7 116.1









#61
1,2-Dibromoethane
Concen: 50.375 ug/l
RT: 11.244 min Scan# 1585
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 248977
Ion Ratio Lower Upper
107 100
109 93.6 74.5 111.7



#62
4-Bromofluorobenzene
Concen: 51.793 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 95 Resp: 405106
Ion Ratio Lower Upper
95 100
174 69.8 0.0 142.6
176 66.5 0.0 134.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 893447
Ion Ratio Lower Upper
117 100
82 57.2 46.3 69.5
119 32.1 24.7 37.1



#64
Tetrachloroethene
Concen: 49.937 ug/l
RT: 10.873 min Scan# 1522
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:164 Resp: 284402
Ion Ratio Lower Upper
164 100
166 126.2 101.8 152.8
129 105.8 78.8 118.2
131 96.4 79.4 119.0









#65
Chlorobenzene
Concen: 51.628 ug/l
RT: 11.667 min Scan# 1657
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 955650
Ion Ratio Lower Upper
112 100
114 32.6 26.5 39.7



#66
1,1,1,2-Tetrachloroethane
Concen: 51.405 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:131 Resp: 342301
Ion Ratio Lower Upper
131 100
133 93.6 46.4 139.1
119 66.8 34.2 102.5









#67
Ethyl Benzene
Concen: 54.679 ug/l
RT: 11.738 min Scan# 1669
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 1781920
Ion Ratio Lower Upper
91 100
106 30.8 24.9 37.3



#68
m/p-Xylenes
Concen: 109.989 ug/l
RT: 11.844 min Scan# 1687
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 106 Resp: 1389518
Ion Ratio Lower Upper
106 100
91 203.3 162.4 243.6









#69
o-Xylene
Concen: 54.582 ug/l
RT: 12.173 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 626873
Ion Ratio Lower Upper
106 100
91 216.8 107.6 322.8



#70
Styrene
Concen: 55.586 ug/l
RT: 12.185 min Scan# 1745
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:104 Resp: 1106368
Ion Ratio Lower Upper
104 100
78 49.5 39.7 59.5
103 54.9 44.0 66.0









#71
Bromoform
Concen: 51.995 ug/l
RT: 12.349 min Scan# 1773
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 168143
Ion Ratio Lower Upper
173 100
175 48.0 24.1 72.3
254 0.0 0.1 0.1#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:152 Resp: 421947
Ion Ratio Lower Upper
152 100
115 62.3 31.1 93.3
150 175.5 0.0 343.0





#73

Isopropylbenzene

Concen: 56.092 ug/l

RT: 12.473 min Scan# 1794

Delta R.T. -0.000 min

Lab File: VD070017.D





Acq: 11 Aug 2021 12:15

Tgt Ion:105 Resp: 1700555

Ion Ratio Lower Upper

105 100

120 25.9 13.0 38.9



#74

N-ethyl acetate

Concen: 58.020 ug/l

RT: 12.279 min Scan# 1761

Delta R.T. -0.000 min

Lab File: VD070017.D

Acq: 11 Aug 2021 12:15

Tgt Ion: 43 Resp: 361812




Ion Ratio Lower Upper


43 100

70 47.3 37.9 56.9

55 27.9 23.7 35.5

61 27.2 20.9 31.3









#75
1,1,2,2-Tetrachloroethane
Concen: 51.536 ug/l
RT: 12.720 min Scan# 1836
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 312024
Ion Ratio Lower Upper
83 100
131 9.2 4.9 14.6
85 64.3 31.8 95.4



#76
1,2,3-Trichloropropane
Concen: 88.533 ug/l
RT: 12.773 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 75 Resp: 397069
Ion Ratio Lower Upper
75 100
77 0.0 181.5 544.5#









#77
Bromobenzene
Concen: 51.073 ug/l
RT: 12.749 min Scan# 1841
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 348738
Ion Ratio Lower Upper
156 100
77 217.6 107.1 321.3
158 96.5 48.1 144.3



#78
n-propylbenzene
Concen: 56.242 ug/l
RT: 12.808 min Scan# 1851
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 91 Resp: 2190013
Ion Ratio Lower Upper
91 100
120 21.7 10.9 32.7









#79
2-Chlorotoluene
Concen: 53.739 ug/l
RT: 12.897 min Scan# 1866
Delta R.T. -0.006 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 1215473
Ion Ratio Lower Upper
91 100
126 31.8 15.9 47.7



#80
1,3,5-Trimethylbenzene
Concen: 55.773 ug/l
RT: 12.949 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:105 Resp: 1446444
Ion Ratio Lower Upper
105 100
120 48.3 23.8 71.2









#81
trans-1,4-Dichloro-2-butene
Concen: 58.194 ug/l
RT: 12.520 min Scan# 1802
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 86033
Ion Ratio Lower Upper
75 100
53 86.6 67.2 100.8
89 45.9 35.8 53.6



#82
4-Chlorotoluene
Concen: 53.107 ug/l
RT: 12.996 min Scan# 1883
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 91 Resp: 1255772
Ion Ratio Lower Upper
91 100
126 31.8 15.9 47.6









#83
tert-Butylbenzene
Concen: 56.724 ug/l
RT: 13.214 min Scan# 1920
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 1196061
Ion Ratio Lower Upper
119 100
91 66.6 33.7 101.1
134 26.0 13.3 39.8



#84
1,2,4-Trimethylbenzene
Concen: 55.675 ug/l
RT: 13.261 min Scan# 1928
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:105 Resp: 1424464
Ion Ratio Lower Upper
105 100
120 44.1 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 57.268 ug/l
 RT: 13.391 min Scan# 1950
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 1908960
 Ion Ratio Lower Upper
 105 100
 134 18.8 9.6 28.6



#86
 p-Isopropyltoluene
 Concen: 56.881 ug/l
 RT: 13.508 min Scan# 1970
 Delta R.T. -0.000 min
 Lab File: VD070017.D
 Acq: 11 Aug 2021 12:15

Tgt Ion:119 Resp: 1540418
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.6 37.8
 91 23.2 11.7 35.1









#87
1,3-Dichlorobenzene
Concen: 51.447 ug/l
RT: 13.508 min Scan# 1970
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 737537
Ion Ratio Lower Upper
146 100
111 43.1 21.7 65.1
148 64.5 32.1 96.3



#88
1,4-Dichlorobenzene
Concen: 50.095 ug/l
RT: 13.585 min Scan# 1983
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:146 Resp: 715794
Ion Ratio Lower Upper
146 100
111 41.8 20.8 62.4
148 64.1 31.6 95.0









#89
n-Butylbenzene
Concen: 57.078 ug/l
RT: 13.832 min Scan# 2025
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion: 91 Resp: 1525488
Ion Ratio Lower Upper
91 100
92 53.9 26.6 79.7
134 24.1 11.9 35.7



#90
Hexachloroethane
Concen: 53.143 ug/l
RT: 14.102 min Scan# 2071
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 117 Resp: 304859
Ion Ratio Lower Upper
117 100
201 62.9 31.9 95.5









#91
1,2-Dichlorobenzene
Concen: 52.055 ug/l
RT: 13.879 min Scan# 2033
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 644093
Ion Ratio Lower Upper
146 100
111 43.0 21.7 65.1
148 62.3 31.2 93.6



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.068 ug/l
RT: 14.496 min Scan# 2138
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion: 75 Resp: 44523
Ion Ratio Lower Upper
75 100
155 73.6 36.3 108.9
157 98.5 47.0 141.0









#93
1,2,4-Trichlorobenzene
Concen: 54.405 ug/l
RT: 15.149 min Scan# 2249
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15




Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 363198
Ion Ratio Lower Upper
180 100
182 94.6 47.6 142.9
145 31.4 16.4 49.2



#94
Hexachlorobutadiene
Concen: 53.501 ug/l
RT: 15.255 min Scan# 2267
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:225 Resp: 218482
Ion Ratio Lower Upper
225 100
223 61.9 31.5 94.5
227 63.5 31.9 95.7









#95
Naphthalene
Concen: 49.944 ug/l
RT: 15.385 min Scan# 2289
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Instrument :
MSVOA_D
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 688721
Ion Ratio Lower Upper
128 100
127 13.1 10.6 16.0
129 10.7 8.6 12.8



#96
1,2,3-Trichlorobenzene
Concen: 54.512 ug/l
RT: 15.579 min Scan# 2322
Delta R.T. -0.000 min
Lab File: VD070017.D
Acq: 11 Aug 2021 12:15

Tgt Ion:180 Resp: 314711
Ion Ratio Lower Upper
180 100
182 94.5 47.3 142.0
145 34.2 17.3 51.9

