

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070023.D
 Acq On : 11 Aug 2021 15:06
 Operator : VA/SY
 Sample : M3333-03RE
 Misc : 4.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 124030RE

Quant Time: Aug 12 05:29:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	48028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	92979	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	93919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	38073	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	49562	89.727	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 179.460%#	
35) Dibromofluoromethane	7.914	113	33434	57.041	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 114.080%	
50) Toluene-d8	10.338	98	117209	51.860	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.720%	
62) 4-Bromofluorobenzene	12.626	95	38354	50.777	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070023.D
Acq On : 11 Aug 2021 15:06
Operator : VA/SY
Sample : M3333-03RE
Misc : 4.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
124030RE

Quant Time: Aug 12 05:29:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

