

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070027.D
 Acq On : 11 Aug 2021 17:01
 Operator : VA/SY
 Sample : M3340-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 205-239-929

Quant Time: Aug 12 05:30:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	741368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1514930	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1433534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	625244	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	391840	45.956	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.920%
35) Dibromofluoromethane	7.914	113	439574	46.028	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.060%
50) Toluene-d8	10.338	98	2002910	54.390	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.780%
62) 4-Bromofluorobenzene	12.626	95	666796	54.181	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	108.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070027.D
Acq On : 11 Aug 2021 17:01
Operator : VA/SY
Sample : M3340-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
205-239-929

Quant Time: Aug 12 05:30:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

