

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070028.D
 Acq On : 11 Aug 2021 17:30
 Operator : VA/SY
 Sample : M3334-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-235

Quant Time: Aug 12 05:31:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	740380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1484538	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1375762	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	595676	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	450249	52.877	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.760%
35) Dibromofluoromethane	7.914	113	476446	50.910	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.820%
50) Toluene-d8	10.338	98	1888487	52.333	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.660%
62) 4-Bromofluorobenzene	12.626	95	614373	50.943	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070028.D
Acq On : 11 Aug 2021 17:30
Operator : VA/SY
Sample : M3334-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-235

Quant Time: Aug 12 05:31:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

