

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070029.D
 Acq On : 11 Aug 2021 17:59
 Operator : VA/SY
 Sample : M3345-02
 Misc : 7.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DRUM-3-18B

Quant Time: Aug 12 05:31:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.98
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.86
63) Chlorobenzene-d5	0.000	117	0m	50.000	ug/l	-11.64
72) 1,4-Dichlorobenzene-d4	0.000	152	0m	50.000	ug/l	-13.57
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0	0.000	ug/l	
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.344	98	61	0.000	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	12.667	95	66	0.000	ug/l	0.04
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070029.D
Acq On : 11 Aug 2021 17:59
Operator : VA/SY
Sample : M3345-02
Misc : 7.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DRUM-3-18B

Quant Time: Aug 12 05:31:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

