

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
 Data File : VD070033.D
 Acq On : 11 Aug 2021 19:54
 Operator : VA/SY
 Sample : M3331-08
 Misc : 4.82G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BSY1-SB-13-0-1

Quant Time: Aug 12 05:32:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 04:51:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	29095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.867	114	51541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	49688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	18575	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	26575	79.419	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	158.840%
35) Dibromofluoromethane	7.914	113	17423	53.623	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.240%
50) Toluene-d8	10.332	98	62228	49.669	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	99.340%
62) 4-Bromofluorobenzene	12.626	95	17178	41.026	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	82.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081121\
Data File : VD070033.D
Acq On : 11 Aug 2021 19:54
Operator : VA/SY
Sample : M3331-08
Misc : 4.82G/5.00ml/MSVOA_D/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BSY1-SB-13-0-1

Quant Time: Aug 12 05:32:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D072221S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 04:51:13 2021
Response via : Initial Calibration

