

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
 Data File : VD074041.D
 Acq On : 11 Aug 2022 14:56
 Operator : VA/SY
 Sample : N4130-05
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB05

Quant Time: Aug 11 18:02:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	104309	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	179767	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	156999	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	58317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	52295	65.929	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	131.860%
35) Dibromofluoromethane	7.908	113	61226	51.278	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.560%
50) Toluene-d8	10.332	98	143369	45.376	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	90.760%
62) 4-Bromofluorobenzene	12.620	95	57767	39.342	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	78.680%
Target Compounds						
16) Acetone	4.150	43	9220	55.242	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
Data File : VD074041.D
Acq On : 11 Aug 2022 14:56
Operator : VA/SY
Sample : N4130-05
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB05

Quant Time: Aug 11 18:02:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

