

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
 Data File : VD074045.D
 Acq On : 11 Aug 2022 16:52
 Operator : VA/SY
 Sample : N4130-13
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB04

Quant Time: Aug 11 18:03:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	26118	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	49067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	50009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	20232	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	26156	140.357	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 280.720%#	
35) Dibromofluoromethane	7.908	113	22440	68.856	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 137.720%	
50) Toluene-d8	10.332	98	39901	46.393	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.780%	
62) 4-Bromofluorobenzene	12.620	95	18781	46.862	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 93.720%	
Target Compounds						
16) Acetone	4.161	43	4238	101.410	ug/l	Qvalue # 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
Data File : VD074045.D
Acq On : 11 Aug 2022 16:52
Operator : VA/SY
Sample : N4130-13
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB04

Quant Time: Aug 11 18:03:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

