

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
 Data File : VD074046.D
 Acq On : 11 Aug 2022 17:21
 Operator : VA/SY
 Sample : N4130-15
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB01

Quant Time: Aug 11 18:03:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	69504	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	125505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	121049	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	50514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	43706	84.901	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 169.800%#		
35) Dibromofluoromethane	7.908	113	45622	54.730	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 109.460%		
50) Toluene-d8	10.332	98	99340	44.986	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 89.980%		
62) 4-Bromofluorobenzene	12.620	95	45794	44.672	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 89.340%		
Target Compounds						
						Qvalue
16) Acetone	4.167	43	7347	66.063	ug/l #	87
20) Methylene Chloride	4.955	84	7319	7.330	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
Data File : VD074046.D
Acq On : 11 Aug 2022 17:21
Operator : VA/SY
Sample : N4130-15
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB01

Quant Time: Aug 11 18:03:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

