

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
 Data File : VD074049.D
 Acq On : 11 Aug 2022 18:49
 Operator : VA/SY
 Sample : N4130-05RE
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB05RE

Quant Time: Aug 12 01:18:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	16787	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	31726	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	33974	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	13389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	17748	148.660	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	297.320%#	
35) Dibromofluoromethane	7.908	113	15094	71.630	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	143.260%	
50) Toluene-d8	10.332	98	26853	48.549	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	97.100%	
62) 4-Bromofluorobenzene	12.614	95	12808	49.426	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	98.860%	
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	6236	232.163	ug/l #	72
20) Methylene Chloride	4.944	84	2372	10.923	ug/l #	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
Data File : VD074049.D
Acq On : 11 Aug 2022 18:49
Operator : VA/SY
Sample : N4130-05RE
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB05RE

Quant Time: Aug 12 01:18:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

