



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081122\
Data File : VD074034.D
Acq On : 11 Aug 2022 09:41
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Title : SW846 8260
Last Update : Wed Aug 10 14:55:58 2022



AutoFind: Scans 1817, 1818, 1819; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.4	9057	PASS
75	95	30	60	54.7	26997	PASS
95	95	100	100	100.0	49352	PASS
96	95	5	9	6.9	3387	PASS
173	174	0.00	2	1.8	704	PASS
174	95	50	100	79.3	39133	PASS
175	174	5	9	8.4	3279	PASS
176	174	95	101	99.1	38781	PASS
177	176	5	9	6.7	2593	PASS