



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081123\
Data File : VD076977.D
Acq On : 11 Aug 2023 08:37
Operator : JC/SY
Sample : BFB156
Misc : 5.00G/10.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB156

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM080323SMA.M
Title : SFAM01.0
Last Update : Sat Aug 12 02:16:48 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1860

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.2	19709	PASS
75	95	30	80	51.0	52424	PASS
95	95	100	100	100.0	102709	PASS
96	95	5	9	6.5	6665	PASS
173	174	0.00	2	0.9	883	PASS
174	95	50	120	90.8	93288	PASS
175	174	5	9	7.7	7209	PASS
176	174	95	105	97.3	90755	PASS
177	176	5	10	7.3	6594	PASS

Scan 1860 (12.528 min): VD076977.D\data.ms
BFB156

m/z	Abundance
40.10	642.0
44.00	531.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

m/z	Abundance
36.10	834.0
37.10	4291.0
38.10	3651.0
39.20	1642.0
40.10	713.0
42.90	155.0
44.00	887.0
45.00	928.0
46.10	232.0
47.10	1250.0
48.00	648.0
49.00	4101.0
50.10	18720.0
51.10	5417.0
52.00	211.0
55.20	437.0
56.00	1438.0
57.10	3290.0
58.10	201.0
59.90	1011.0
61.00	5514.0
62.10	4843.0
63.00	4444.0
64.10	760.0
65.00	736.0
66.90	259.0
68.10	10502.0
69.10	10502.0
70.10	815.0
72.00	594.0
73.00	5927.0
74.10	19240.0
75.10	48952.0
76.10	4143.0
76.90	743.0
78.20	726.0
79.00	5271.0
80.00	1153.0
81.00	5015.0
82.00	1190.0
87.00	5629.0
88.10	4727.0
90.90	405.0
92.00	3178.0
93.00	5155.0
94.00	12754.0
95.10	94728.0
96.10	6121.0
102.90	210.0
103.90	745.0
105.00	256.0
105.90	497.0
109.90	194.0
110.80	163.0
112.70	256.0
116.00	729.0
116.90	1153.0
118.00	548.0
119.00	725.0
124.90	211.0
128.00	765.0
129.00	299.0
130.00	373.0
130.90	171.0
133.20	174.0
134.80	372.0
136.80	252.0
141.00	1148.0
142.10	284.0
142.90	1402.0
144.10	184.0
144.90	510.0
147.90	414.0
155.10	199.0
156.90	188.0
158.90	169.0
160.90	332.0
170.90	283.0
172.20	501.0
172.90	1041.0
174.00	80400.0
175.00	6570.0
176.00	81160.0
177.00	5596.0
178.10	175.0
191.00	317.0
193.10	372.0
207.20	193.0
265.00	221.0
265.40	189.0
281.10	1177.0
282.10	483.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

283.30

382.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

m/z	Abundance
36.00	1149.0
37.00	5120.0
38.10	4618.0
39.00	2234.0
40.10	995.0
44.10	1553.0
45.00	916.0
46.20	250.0
47.20	1299.0
47.80	407.0
49.10	5228.0
50.10	21664.0
51.10	6610.0
52.00	293.0
55.00	455.0
56.10	1934.0
57.00	3177.0
59.90	1277.0
61.10	5957.0
62.10	5186.0
63.10	5112.0
64.00	681.0
65.10	599.0
67.10	328.0
68.10	13104.0
69.10	11885.0
70.20	1081.0
71.90	689.0
73.10	6416.0
74.10	23688.0
75.10	58016.0
76.10	5375.0
77.10	770.0
77.90	650.0
79.00	5894.0
80.00	1550.0
81.00	6607.0
81.90	1041.0
83.00	361.0
85.70	163.0
87.10	6216.0
88.10	6083.0
91.00	758.0
91.90	4158.0
93.10	5595.0
94.10	14631.0
95.10	116104.0
96.10	6910.0
97.00	374.0
102.80	211.0
103.70	416.0
104.80	227.0
105.90	794.0
110.30	212.0
110.80	238.0
112.10	173.0
112.90	200.0
116.10	642.0
117.00	1002.0
117.70	407.0
118.10	617.0
118.90	820.0
127.90	358.0
128.80	258.0
129.80	519.0
132.90	535.0
133.80	163.0
135.00	366.0
137.00	253.0
141.00	1540.0
141.80	251.0
142.90	1986.0
144.80	461.0
146.10	290.0
147.80	365.0
153.10	155.0
155.10	372.0
156.90	210.0
158.90	222.0
160.80	221.0
170.30	210.0
170.80	184.0
171.10	194.0
171.90	795.0
173.00	811.0
174.00	104496.0
175.10	8374.0
176.00	97360.0
177.00	7817.0
177.90	313.0
178.80	187.0
191.00	248.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

192.90	227.0
207.00	278.0
249.00	182.0
265.30	315.0
281.10	1989.0
282.10	615.0
283.20	356.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

m/z	Abundance
36.00	1186.0
37.10	3916.0
38.10	3714.0
39.10	1823.0
40.00	1077.0
43.90	1245.0
45.10	906.0
47.10	1333.0
48.10	691.0
49.10	4640.0
50.10	18744.0
51.10	6199.0
52.10	312.0
55.00	453.0
56.10	1361.0
57.10	3229.0
60.00	1334.0
61.00	5116.0
62.00	4675.0
63.00	4782.0
64.00	736.0
64.90	296.0
66.90	240.0
68.00	9063.0
69.00	9397.0
70.00	743.0
72.00	346.0
73.10	6129.0
74.10	18272.0
75.10	50304.0
76.10	4656.0
77.00	597.0
78.00	453.0
79.00	5753.0
79.90	1077.0
81.00	4964.0
81.90	1259.0
83.00	284.0
87.00	4880.0
88.00	4288.0
91.10	527.0
92.00	2843.0
93.10	4757.0
94.10	14114.0
95.10	97296.0
96.10	6965.0
103.10	256.0
103.90	677.0
105.10	200.0
105.90	427.0
107.90	170.0
110.90	219.0
114.90	178.0
115.80	452.0
116.90	1135.0
118.10	437.0
119.00	970.0
123.90	150.0
124.80	270.0
125.60	167.0
127.80	321.0
128.00	327.0
129.00	281.0
130.00	292.0
133.00	541.0
133.80	210.0
134.90	279.0
136.80	237.0
141.00	1415.0
142.10	160.0
143.00	1701.0
144.90	249.0
146.80	250.0
147.80	231.0
153.10	206.0
155.00	408.0
161.00	169.0
162.90	155.0
170.90	169.0
172.10	904.0
173.10	797.0
174.00	94968.0
175.10	6684.0
176.00	93744.0
177.00	6369.0
179.00	163.0
191.10	267.0
193.10	685.0
207.10	496.0
249.00	236.0
265.20	451.0
281.10	2778.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156

282.10855.0

283.10538.0

Instrument :
MSVOA_D
ClientSampleId :
BFB156