

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081219\
 Data File : VD063488.D
 Acq On : 12 Aug 2019 15:32
 Operator : JC/SY
 Sample : K4245-04
 Misc : 6.82g/5ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC-26

Quant Time: Aug 13 08:22:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	682862	50.00	ug/l	-0.02
34) 1,4-Difluorobenzene	8.22	114	1023176	50.00	ug/l	-0.02
63) Chlorobenzene-d5	12.36	117	508923	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	14.52	152	110379	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.45	65	362688	49.69	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	6.92	113	839	0.10	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	0.20%	
50) Toluene-d8	10.41	98	654430	34.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	69.08%	
62) 4-Bromofluorobenzene	13.55	95	151408	17.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	35.26%	
Target Compounds						
16) Acetone	3.23	43	69911	47.116	ug/l	99
20) Methylene Chloride	3.82	84	41067	6.401	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081219\
Data File : VD063488.D
Acq On : 12 Aug 2019 15:32
Operator : JC/SY
Sample : K4245-04
Misc : 6.82g/5ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC-26

Quant Time: Aug 13 08:22:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

