

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD081219\
 Data File : VD063516.D
 Acq On : 13 Aug 2019 6:01
 Operator : JC/SY
 Sample : K4223-30RE
 Misc : 4.32g/5ml/MSVOA D/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-15RE

Quant Time: Aug 13 09:34:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 10 06:42:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.05	168	721682	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	8.23	114	999830	50.00	ug/l	0.00
63) Chlorobenzene-d5	12.37	117	343251	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	14.52	152	198876	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.46	65	356558	46.22	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	92.44%	
35) Dibromofluoromethane	6.93	113	158332	18.80	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	37.60%	
50) Toluene-d8	10.41	98	194887	10.53	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	21.06%	
62) 4-Bromofluorobenzene	13.55	95	164510	19.61	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	39.22%	
Target Compounds						
16) Acetone	3.23	43	155835	99.375	ug/l	95
20) Methylene Chloride	3.83	84	62295	9.187	ug/l #	93
25) 2-Butanone	6.08	43	44665	28.839	ug/l #	89
29) Tetrahydrofuran	6.46	42	9405	13.032	ug/l	97
52) Toluene	10.51	92	14915	1.191	ug/l	99
86) p-Isopropyltoluene	14.49	119	34055	2.882	ug/l	96
95) Naphthalene	16.12	128	12334	2.085	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD081219\
Data File : VD063516.D
Acq On : 13 Aug 2019 6:01
Operator : JC/SY
Sample : K4223-30RE
Misc : 4.32g/5ml/MSVOA D/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-15RE

Quant Time: Aug 13 09:34:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 10 06:42:47 2019
Response via : Initial Calibration

