


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD081219\
Data File : VD063477.D
Acq On : 12 Aug 2019 9:29
Operator : JC/SY
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D080919S.M
Title : SW846 8260
Last Update : Sat Aug 10 06:42:47 2019



AutoFind: Scans 1231, 1232, 1233; Background Corrected with Scan 1227

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	58613	PASS
75	95	30	60	44.5	123672	PASS
95	95	100	100	100.0	278208	PASS
96	95	5	9	6.8	18945	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.3	240018	PASS
175	174	5	9	7.3	17525	PASS
176	174	95	101	98.6	236616	PASS
177	176	5	9	6.0	14285	PASS