

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074059.D
 Acq On : 12 Aug 2022 12:51
 Operator : VA/SY
 Sample : N4119-04RE
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP2-2FTRE

Quant Time: Aug 12 13:24:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	146572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	245311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	210213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	79114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	55615	47.787	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.580%
35) Dibromofluoromethane	7.902	113	69318	42.544	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.080%
50) Toluene-d8	10.332	98	159004	35.682	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	71.360%
62) 4-Bromofluorobenzene	12.620	95	61782	30.834	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	61.660%
Target Compounds						
16) Acetone	4.150	43	3658m	15.597	ug/l	
81) trans-1,4-Dichloro-2-b...	12.614	75	34001	125.634	ug/l #	13

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074059.D
Acq On : 12 Aug 2022 12:51
Operator : VA/SY
Sample : N4119-04RE
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP2-2FTRE

Quant Time: Aug 12 13:24:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

