

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074060.D
 Acq On : 12 Aug 2022 13:20
 Operator : VA/SY
 Sample : N4127-01RE
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-080922RE

Quant Time: Aug 12 13:37:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	149512	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.856	114	265203	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	242880	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	103042	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	55698	46.759	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.520%
35) Dibromofluoromethane	7.909	113	66610	37.815	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.640%
50) Toluene-d8	10.332	98	162700	33.431	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	66.860%
62) 4-Bromofluorobenzene	12.620	95	73927	34.128	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	68.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074060.D
Acq On : 12 Aug 2022 13:20
Operator : VA/SY
Sample : N4127-01RE
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-080922RE

Quant Time: Aug 12 13:37:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

