

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074062.D
 Acq On : 12 Aug 2022 14:51
 Operator : VA/SY
 Sample : N4107-07
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CONCRETE-STOCKPILE

Quant Time: Aug 12 23:47:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	164032	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	280507	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	249603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	108168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	57445	43.436	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	86.880%
35) Dibromofluoromethane	7.903	113	69115	37.097	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.200%
50) Toluene-d8	10.332	98	173413	33.737	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	67.480%
62) 4-Bromofluorobenzene	12.620	95	75224	32.833	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074062.D
Acq On : 12 Aug 2022 14:51
Operator : VA/SY
Sample : N4107-07
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CONCRETE-STOCKPILE

Quant Time: Aug 12 23:47:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

