

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074065.D
 Acq On : 12 Aug 2022 16:19
 Operator : VA/SY
 Sample : N4121-02
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Aug 12 23:48:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	135862	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	233791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	204146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	82714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	55089	51.662	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.320%
35) Dibromofluoromethane	7.908	113	63324	40.780	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	81.560%
50) Toluene-d8	10.326	98	147067	34.441	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	68.880%
62) 4-Bromofluorobenzene	12.620	95	63798	33.410	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.820%
Target Compounds						
16) Acetone	4.167	43	3446	15.852	ug/l	Qvalue # 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074065.D
Acq On : 12 Aug 2022 16:19
Operator : VA/SY
Sample : N4121-02
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Aug 12 23:48:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

