

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074067.D
 Acq On : 12 Aug 2022 17:18
 Operator : VA/SY
 Sample : N4183-03
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 STOCKPILE

Quant Time: Aug 12 23:48:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	13567	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	24901	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	26531	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	12936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	9924	100.179	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	200.360%	#
35) Dibromofluoromethane	7.908	113	8514	51.478	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	102.960%	
50) Toluene-d8	10.326	98	15712	34.566	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	69.140%	
62) 4-Bromofluorobenzene	12.614	95	8142	40.032	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	80.060%	
Target Compounds						
						Qvalue
16) Acetone	4.168	43	2494	114.888	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074067.D
Acq On : 12 Aug 2022 17:18
Operator : VA/SY
Sample : N4183-03
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
STOCKPILE

Quant Time: Aug 12 23:48:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

