

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074068.D
 Acq On : 12 Aug 2022 17:47
 Operator : VA/SY
 Sample : N4153-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-081022

Quant Time: Aug 12 23:48:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	115591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	209981	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	194302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	86008	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	47689	52.717	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	105.440%	
35) Dibromofluoromethane	7.908	113	54686	39.211	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	78.420%	
50) Toluene-d8	10.332	98	118804	30.334	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	60.660%	
62) 4-Bromofluorobenzene	12.620	95	55650	32.447	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	64.900%	
Target Compounds						
					Qvalue	
5) Bromomethane	2.773	94	1755	1.035	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074068.D
Acq On : 12 Aug 2022 17:47
Operator : VA/SY
Sample : N4153-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-081022

Quant Time: Aug 12 23:48:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

