

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074071.D
 Acq On : 12 Aug 2022 19:14
 Operator : VA/SY
 Sample : N4140-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 LINDEN-MR-B1B-SP

Quant Time: Aug 12 23:48:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	143494	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.855	114	260927	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	226836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	95297	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	51899	45.144	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.280%
35) Dibromofluoromethane	7.903	113	65461	37.772	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	75.540%
50) Toluene-d8	10.326	98	149496	30.799	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	61.600%
62) 4-Bromofluorobenzene	12.620	95	63859	29.964	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	59.920%
Target Compounds						
					Qvalue	
16) Acetone	4.156	43	3254	14.172	ug/l	98
17) Carbon Disulfide	4.397	76	8499	1.827	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074071.D
Acq On : 12 Aug 2022 19:14
Operator : VA/SY
Sample : N4140-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
LINDEN-MR-B1B-SP

Quant Time: Aug 12 23:48:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

