

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
 Data File : VD074074.D
 Acq On : 12 Aug 2022 20:41
 Operator : VA/SY
 Sample : N4136-03
 Misc : 10.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 WC-5

Quant Time: Aug 12 23:49:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 10 14:55:58 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.961	168	3173	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.855	114	5563	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.632	117	6009	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	2481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	2110	90.283	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	180.560%	#
35) Dibromofluoromethane	7.914	113	1773	47.985	ug/l	0.01
Spiked Amount	50.000	Range 54 - 147	Recovery	=	95.980%	
50) Toluene-d8	10.332	98	3253	31.566	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	63.140%	
62) 4-Bromofluorobenzene	12.620	95	2257	49.672	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	99.340%	
Target Compounds						
16) Acetone	4.144	43	2079	409.492	ug/l	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081222\
Data File : VD074074.D
Acq On : 12 Aug 2022 20:41
Operator : VA/SY
Sample : N4136-03
Misc : 10.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
WC-5

Quant Time: Aug 12 23:49:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081022S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 10 14:55:58 2022
Response via : Initial Calibration

