

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
 Data File : VD070059.D
 Acq On : 13 Aug 2021 19:38
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 02:52:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	0.572	0.419	26.7#	77	0.00
3 P	Chloromethane	0.785	0.712	9.3	89	0.00
4 C	Vinyl Chloride	0.846	0.720	14.9#	81	0.00
5 T	Bromomethane	0.532	0.468	12.0	88	0.00
6 T	Chloroethane	0.541	0.494	8.7	86	0.00
7 T	Trichlorofluoromethane	1.016	0.866	14.8	80	0.00
8 T	Diethyl Ether	0.291	0.304	-4.5	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.647	0.574	11.3	84	0.00
10 T	Methyl Iodide	0.591	0.513	13.2	70	0.00
11 T	Tert butyl alcohol	0.033	0.029	12.1	84	0.00
12 CM	1,1-Dichloroethene	0.594	0.547	7.9#	83	0.00
13 T	Acrolein	0.037	0.032	13.5	75	0.00
14 T	Allyl chloride	0.886	0.843	4.9	81	0.00
15 T	Acrylonitrile	0.135	0.140	-3.7	89	0.00
16 T	Acetone	0.108	0.103	4.6	84	0.00
17 T	Carbon Disulfide	2.148	1.963	8.6	83	0.00
18 T	Methyl Acetate	0.323	0.322	0.3	86	0.00
19 T	Methyl tert-butyl Ether	1.233	1.297	-5.2	87	0.00
20 T	Methylene Chloride	0.968	0.787	18.7	94	0.00
21 T	trans-1,2-Dichloroethene	0.690	0.692	-0.3	88	0.00
22 T	Diisopropyl ether	1.864	2.012	-7.9	88	0.00
23 T	Vinyl Acetate	0.845	0.921	-9.0	86	0.00
24 P	1,1-Dichloroethane	1.275	1.258	1.3	88	0.00
25 T	2-Butanone	0.148	0.150	-1.4	85	0.00
26 T	2,2-Dichloropropane	0.993	0.867	12.7	77	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.760	-3.7	90	0.00
28 T	Bromochloromethane	0.464	0.457	1.5	93	0.00
29 T	Tetrahydrofuran	0.097	0.103	-6.2	88	0.00
30 C	Chloroform	1.254	1.261	-0.6#	90	0.00
31 T	Cyclohexane	1.164	1.028	11.7	80	0.00
32 T	1,1,1-Trichloroethane	1.048	0.999	4.7	85	0.00
33 S	1,2-Dichloroethane-d4	0.588	0.595	-1.2	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.323	0.325	-0.6	90	0.00
36 T	1,1-Dichloropropene	0.538	0.502	6.7	83	0.00
37 T	Ethyl Acetate	0.195	0.203	-4.1	86	0.00
38 T	Carbon Tetrachloride	0.488	0.448	8.2	82	0.00
39 T	Methylcyclohexane	0.626	0.580	7.3	80	0.00
40 TM	Benzene	1.560	1.559	0.1	88	0.00
41 T	Methacrylonitrile	0.106	0.104	1.9	89	0.00
42 TM	1,2-Dichloroethane	0.400	0.406	-1.5	91	0.00
43 T	Isopropyl Acetate	0.365	0.364	0.3	87	0.00
44 TM	Trichloroethene	0.377	0.366	2.9	88	0.00
45 C	1,2-Dichloropropane	0.405	0.410	-1.2#	90	0.00
46 T	Dibromomethane	0.197	0.203	-3.0	92	0.00
47 T	Bromodichloromethane	0.509	0.521	-2.4	91	0.00
48 T	Methyl methacrylate	0.174	0.185	-6.3	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
 Data File : VD070059.D
 Acq On : 13 Aug 2021 19:38
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 14 02:52:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	91	0.00
50 S	Toluene-d8	1.256	1.287	-2.5	89	0.00
51 T	4-Methyl-2-Pentanone	0.190	0.201	-5.8	88	0.00
52 CM	Toluene	0.950	0.970	-2.1#	88	0.00
53 T	t-1,3-Dichloropropene	0.451	0.432	4.2	83	0.00
54 T	cis-1,3-Dichloropropene	0.547	0.512	6.4	83	0.00
55 T	1,1,2-Trichloroethane	0.279	0.290	-3.9	92	0.00
56 T	Ethyl methacrylate	0.317	0.341	-7.6	88	0.00
57 T	1,3-Dichloropropane	0.483	0.495	-2.5	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.167	0.176	-5.4	82	0.00
59 T	2-Hexanone	0.127	0.134	-5.5	86	0.00
60 T	Dibromochloromethane	0.319	0.329	-3.1	92	0.00
61 T	1,2-Dibromoethane	0.252	0.261	-3.6	90	0.00
62 S	4-Bromofluorobenzene	0.414	0.430	-3.9	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.323	0.305	5.6	88	0.00
65 PM	Chlorobenzene	1.045	1.029	1.5	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.365	1.6	90	0.00
67 C	Ethyl Benzene	1.847	1.848	-0.1#	87	0.00
68 T	m/p-Xylenes	0.716	0.725	-1.3	89	0.00
69 T	o-Xylene	0.646	0.664	-2.8	89	0.00
70 T	Styrene	1.131	1.196	-5.7	90	0.00
71 P	Bromoform	0.181	0.181	0.0	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
73 T	Isopropylbenzene	3.646	3.543	2.8	87	0.00
74 T	N-amyl acetate	0.786	0.764	2.8	86	0.00
75 P	1,1,2,2-Tetrachloroethane	0.728	0.711	2.3	93	0.00
76 T	1,2,3-Trichloropropane	0.507	0.476	6.1	93	0.00
77 T	Bromobenzene	0.809	0.810	-0.1	92	0.00
78 T	n-propylbenzene	4.702	4.570	2.8	86	0.00
79 T	2-Chlorotoluene	2.705	2.614	3.4	88	0.00
80 T	1,3,5-Trimethylbenzene	3.104	3.056	1.5	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.195	0.173	11.3	76	0.00
82 T	4-Chlorotoluene	2.833	2.770	2.2	89	0.00
83 T	tert-Butylbenzene	2.523	2.450	2.9	87	0.00
84 T	1,2,4-Trimethylbenzene	3.079	3.086	-0.2	88	0.00
85 T	sec-Butylbenzene	4.042	3.835	5.1	85	0.00
86 T	p-Isopropyltoluene	3.266	3.171	2.9	85	0.00
87 T	1,3-Dichlorobenzene	1.719	1.649	4.1	90	0.00
88 T	1,4-Dichlorobenzene	1.703	1.619	4.9	90	0.00
89 T	n-Butylbenzene	3.205	3.024	5.6	83	0.00
90 T	Hexachloroethane	0.689	0.633	8.1	88	0.00
91 T	1,2-Dichlorobenzene	1.473	1.439	2.3	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.109	0.099	9.2	87	0.00
93 T	1,2,4-Trichlorobenzene	0.819	0.786	4.0	87	0.00
94 T	Hexachlorobutadiene	0.484	0.447	7.6	86	0.00
95 T	Naphthalene	1.491	1.508	-1.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
Data File : VD070059.D
Acq On : 13 Aug 2021 19:38
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Aug 14 02:52:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.715	0.708	1.0	89	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6