

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
 Data File : VD070058.D
 Acq On : 13 Aug 2021 19:09
 Operator : VA/SY
 Sample : M3368-15
 Misc : 9.84G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SWCS-06-1015

Quant Time: Aug 14 02:52:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	573615	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1148751	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1086025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	457062	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	389097	57.673	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.340%
35) Dibromofluoromethane	7.914	113	381832	51.448	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.900%
50) Toluene-d8	10.338	98	1464057	50.719	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.440%
62) 4-Bromofluorobenzene	12.626	95	450268	47.358	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.720%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
Data File : VD070058.D
Acq On : 13 Aug 2021 19:09
Operator : VA/SY
Sample : M3368-15
Misc : 9.84G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SWCS-06-1015

Quant Time: Aug 14 02:52:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.979 min Scan# 1030
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09




Instrument :
MSVOA_D
ClientSampleId :
SWCS-06-1015


Tgt Ion:168 Resp: 573615
Ion Ratio Lower Upper
168 100
99 58.2 50.3 75.5



#33
1,2-Dichloroethane-d4
Concen: 57.673 ug/l
RT: 8.326 min Scan# 1089
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09

Tgt Ion: 65 Resp: 389097
Ion Ratio Lower Upper
65 100
67 54.0 0.0 107.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09




Instrument :
MSVOA_D
ClientSampleId :
SWCS-06-1015


Tgt Ion:114 Resp: 1148751
Ion Ratio Lower Upper
114 100
63 21.3 0.0 43.6
88 15.2 0.0 33.4



#35
Dibromofluoromethane
Concen: 51.448 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09

Tgt Ion:113 Resp: 381832
Ion Ratio Lower Upper
113 100
111 102.7 81.8 122.6
192 17.0 13.2 19.8









#50
Toluene-d8
Concen: 50.719 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. 0.006 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09




Instrument :
MSVOA_D
ClientSampleId :
SWCS-06-1015


Tgt Ion: 98 Resp: 1464057
Ion Ratio Lower Upper
98 100
100 64.5 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 47.358 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09

Tgt Ion: 95 Resp: 450268
Ion Ratio Lower Upper
95 100
174 75.9 0.0 141.6
176 71.8 0.0 138.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 1652
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09

Instrument :
MSVOA_D
ClientSampleId :
SWCS-06-1015

Tgt Ion:117 Resp: 1086025
Ion Ratio Lower Upper
117 100
82 56.7 47.5 71.3
119 32.5 25.1 37.7



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.000 min
Lab File: VD070058.D
Acq: 13 Aug 2021 19:09

Tgt Ion:152 Resp: 457062
Ion Ratio Lower Upper
152 100
115 61.1 31.2 93.6
150 155.9 0.0 352.0

