



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081321\
Data File : VD070039.D
Acq On : 13 Aug 2021 09:05
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Title : SW846 8260
Last Update : Sat Aug 14 02:33:51 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	17.5	80771	PASS
75	95	30	60	48.7	224896	PASS
95	95	100	100	100.0	462144	PASS
96	95	5	9	6.4	29728	PASS
173	174	0.00	2	1.2	4071	PASS
174	95	50	100	72.1	333077	PASS
175	174	5	9	7.6	25256	PASS
176	174	95	101	96.3	320683	PASS
177	176	5	9	6.1	19541	PASS