

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081324\
 Data File : VD079724.D
 Acq On : 13 Aug 2024 18:27
 Operator : RP/MD
 Sample : P3526-03RE
 Misc : 6.75G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SP-1-GRABRE

Quant Time: Aug 14 05:54:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081224S.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 13 01:06:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	98008	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	194940	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	126753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	24967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	78196	69.350	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.700%
35) Dibromofluoromethane	7.805	113	80474	66.963	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	133.920%
50) Toluene-d8	10.269	98	245604	48.898	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.800%
62) 4-Bromofluorobenzene	12.569	95	34630	24.624	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	49.240%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.793	84	14195	3.836	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081324\
Data File : VD079724.D
Acq On : 13 Aug 2024 18:27
Operator : RP/MD
Sample : P3526-03RE
Misc : 6.75G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SP-1-GRABRE

Quant Time: Aug 14 05:54:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081224S.M
Quant Title : SW846 8260
QLast Update : Tue Aug 13 01:06:11 2024
Response via : Initial Calibration

