

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
 Data File : VD076992.D
 Acq On : 14 Aug 2023 11:08
 Operator : JC/SY
 Sample : VD0814SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0814SBL01

Quant Time: Aug 14 12:44:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 09 07:25:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	219466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	436389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	423682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	186878	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	88949	51.378	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.760%
35) Dibromofluoromethane	7.804	113	111509	47.530	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.060%
50) Toluene-d8	10.269	98	412542	54.748	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.500%
62) 4-Bromofluorobenzene	12.575	95	124523	47.270	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
Data File : VD076992.D
Acq On : 14 Aug 2023 11:08
Operator : JC/SY
Sample : VD0814SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0814SBL01

Quant Time: Aug 14 12:44:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 09 07:25:36 2023
Response via : Initial Calibration

