

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
 Data File : VD077007.D
 Acq On : 14 Aug 2023 18:04
 Operator : JC/SY
 Sample : 03973-03
 Misc : 5.01G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-02-16.5-17.0

Quant Time: Aug 15 04:50:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
 Quant Title : SW846 8260
 QLast Update : Wed Aug 09 07:25:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	17681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	35566	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	39449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	19955	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	11474	82.264	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	164.520%#
35) Dibromofluoromethane	7.810	113	11990	62.707	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.420%
50) Toluene-d8	10.269	98	38248	62.279	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	124.560%
62) 4-Bromofluorobenzene	12.575	95	15120	70.426	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	140.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD081423\
Data File : VD077007.D
Acq On : 14 Aug 2023 18:04
Operator : JC/SY
Sample : 03973-03
Misc : 5.01G/5.0ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-02-16.5-17.0

Quant Time: Aug 15 04:50:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D080823S.M
Quant Title : SW846 8260
QLast Update : Wed Aug 09 07:25:36 2023
Response via : Initial Calibration

